



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 



6042 Cornerstone Court W, Ste B
San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

Data Sheet
MLL4 Complex Chemiluminescent Assay Kit
Catalog #:79757
Size: 96 reactions

DESCRIPTION: The *MLL4 Complex Chemiluminescent Assay Kit* is designed to measure MLL4 activity for screening and profiling applications, using purified MLL4 and its complex components: WDR5, RbBP5, Ash2, and DPY30. The *MLL4 Complex Chemiluminescent Assay Kit* comes in a convenient format, with 8-well strips pre-coated with histone H3 peptide substrate, an antibody against methylated lysine on Histone H3, a secondary HRP-labeled antibody, S-adenosylmethionine, methyltransferase assay buffer, and enough purified MLL4 enzyme complex for 100 enzyme reactions. The key to the *MLL4 Complex Chemiluminescent Assay Kit* is a highly specific antibody that recognizes methylated K4 residue of Histone H3. With this kit, only three simple steps on a microtiter plate are required for methyltransferase detection. First, S-adenosylmethionine is incubated with a sample containing assay buffer and methyltransferase enzyme. Next, primary antibody is added. Finally, the plates are treated with an HRP-labeled secondary antibody followed by the addition of the HRP substrate to produce chemiluminescence that can be measured using a chemiluminescence reader.

COMPONENTS:

Catalog #	Component	Amount	Storage	
100324	MLL4/WDR5/Ash2L/RbBP5/DPY30	3 µg	-80°C	<i>Avoid freeze/ thaw cycles!</i>
52120	20 µM S-adenosylmethionine	250 µl	-80°C	
52140Z	Primary antibody 26	12.5 µl	-80°C	
52160	4x HMT assay buffer 1*	3 ml	-20°C	
52131H	Secondary HRP-labeled antibody 2	10 µl	-80°C	
52100	Blocking buffer	50 ml	+4°C	
79670	ELISA ECL Substrate (2 components)	6 ml each	+4°C	
	96-well strip well plate pre-coated with histone substrate	1 plate	+4°C	

*Add 125 µl of 0.5 M DTT before use.

MATERIALS REQUIRED BUT NOT SUPPLIED:

TBST buffer (1 x TBS, pH 8.0, containing 0.05% Tween-20)
Luminometer or fluorescent microplate reader capable of reading chemiluminescence
Adjustable micropipettor and sterile tips
Rotating or rocker platform
0.5 M DTT (dithiothreitol)

APPLICATIONS: Great for studying enzyme kinetics and HTS applications.

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** or Fax **1.858.481.8694**
Or you can Email us at: info@bpsbioscience.com
Please visit our website at: www.bpsbioscience.com



6042 Cornerstone Court W, Ste B
San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

STABILITY: Up to one year from date of receipt when stored as directed.

REFERENCE(S): Dillon, S.C., *et al. Genome Biology* 2005; **6**:227.

ASSAY PROTOCOL:

Step 1:

- 1) Rehydrate the microwells by adding 200 μ l of TBST buffer (1x TBS, pH 8.0, containing 0.05% Tween-20) to every well. Incubate 15 minutes at room temperature. Tap the strip plate onto clean paper towels to remove liquid.
- 2) Thaw **S-adenosylmethionine** on ice. Upon first thaw, briefly spin tube containing **S-adenosylmethionine** to recover full contents of the tube. Aliquot **S-adenosylmethionine** into single use aliquots and store at -80°C . *Note: S-adenosylmethionine is very sensitive to freeze/thaw cycles. Avoid multiple freeze/thaw cycles.*
- 3) Add 125 μ l of 0.5 M DTT to 4x HMT assay buffer 1. Prepare the master mixture: N wells \times (7.5 μ l **4x HMT assay buffer 1** + 2.5 μ l **20 μ M S-adenosylmethionine** + 15 μ l distilled water). Add 25 μ l of master mixture to all wells labeled "Positive Control", "Test Sample" and "Blank". For wells labeled "Substrate control", add 7.5 μ l **4x HMT assay buffer 1** + 17.5 μ l water.

	Blank	Substrate Control	Positive Control	Test Sample
4x HMT assay buffer 1	7.5 μ l	7.5 μ l	7.5 μ l	7.5 μ l
20 μ M S-adenosylmethionine	2.5 μ l	-	2.5 μ l	2.5 μ l
Distilled water	15 μ l	17.5 μ l	15 μ l	15 μ l
Test Inhibitor	-	-	-	5 μ l
Inhibitor buffer (no inhibitor)	5 μ l	5 μ l	5 μ l	-
1x HMT assay buffer 1	20 μ l	-	-	-
Diluted MLL4 (1.5 ng/ μ l)	-	20 μ l	20 μ l	20 μ l
Total	50 μl	50 μl	50 μl	50 μl

- 4) Add 5 μ l of inhibitor solution to each well designated "Test Inhibitor".
- 5) For the "Positive Control", "Substrate Control" and "Blank", add 5 μ l of the same solution without inhibitor (inhibitor buffer). If test inhibitors are dissolved in DMSO, be sure to keep final DMSO concentration in the assay at $\leq 1\%$.
- 6) Thaw **MLL4 enzyme** on ice. Upon first thaw, briefly spin tube containing enzyme to recover full contents of the tube. Aliquot **MLL4 enzyme** into single use aliquots. Store

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** or Fax **1.858.481.8694**
Or you can Email us at: info@bpsbioscience.com
Please visit our website at: www.bpsbioscience.com



6042 Cornerstone Court W, Ste B
San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

remaining undiluted enzyme in aliquots at -80°C. *Note: **MLL4 enzyme** is very sensitive to freeze/thaw cycles. Do not re-use thawed aliquots or diluted enzyme.*

- 7) Dilute **MLL4 enzyme** in **1x HMT assay buffer 1** to 1.5 ng/μl (30 ng/20 μl). Keep diluted enzyme on ice until use. Discard any unused diluted enzyme after use. *Note: Diluted enzyme may not be stable. Dilute the enzyme immediately before use.*
- 8) Add 20 μl of **1x HMT assay buffer 1** to the wells designated "Blank".
- 9) Initiate reaction by adding 20 μl of diluted **MLL4 enzyme** to the wells designated "Positive Control", "Substrate Control", and "Test Sample ". Incubate at room temperature for one hour.
- 10) Remove the supernatant from the wells and wash the strip three times with 200 μl TBST buffer. Blot dry onto clean paper towels
- 11) Add 100 μl of **Blocking buffer** to every well. Shake on a rotating platform for 10 minutes. Remove supernatant as described above.

Step 2:

- 1) Dilute "**Primary antibody 26**" 800-fold with **Blocking buffer**.
- 2) Add 100 μl per well. Incubate 1 hour at room temperature with slow shaking.
- 3) Wash plate three times with TBST buffer and incubate in **Blocking buffer** as in steps 1-10 and 1-11.

Step 3:

- 1) Dilute "**Secondary HRP-labeled antibody 2**" 1,000-fold with **Blocking buffer**.
- 2) Add 100 μl per well. Incubate for 30 min. at room temperature with slow shaking.
- 3) Wash plate with TBST buffer and incubate in **Blocking buffer** as in steps 1-10 and 1-11.
- 4) Just before use, mix on ice 50 μl **ELISA ECL substrate A** and 50 μl **ELISA ECL substrate B**. Add 100 μl per well. Discard any unused chemiluminescent reagent after use.
- 5) Immediately read sample in a luminometer or microtiter-plate capable of reading chemiluminescence. Blank value is subtracted from all other values.

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.


To place your order, please contact us by Phone **1.858.202.1401** or Fax **1.858.481.8694**
Or you can Email us at: info@bpsbioscience.com
Please visit our website at: www.bpsbioscience.com

Reading Chemiluminescence:

Chemiluminescence is the emission of light (luminescence) which results from a chemical reaction. The detection of chemiluminescence requires no wavelength selection because the method used is emission photometry and is not emission spectrophotometry.

To properly read chemiluminescence, make sure you are using your plate reader in a LUMINESCENCE mode. Typical integration time is 1 second, delay after plate movement is 100 msec. Make sure you don't have filter when emit the light (Synergy 2 BioTek: use "hole" position on filter wheel). Optics position – Top. Read type: endpoint. Sensitivity may be adjusted based on luminescence of a control without enzyme (typically we set this value as 100 when using Synergy 2 plate reader).

Example of Assay Results:



MLL4 complex enzyme activity, measured using the *MLL4 Complex Chemiluminescent Assay Kit*, BPS Bioscience #79757. Luminescence was measured using a Bio-Tek fluorescent microplate reader. *Data shown is lot-specific. For lot-specific information, please contact BPS Bioscience, Inc. at info@bpsbioscience.com*

RELATED PRODUCTS

<u>Product Name</u>	<u>Catalog #</u>	<u>Size</u>
MLL1/WDR5/Ash2L/RbBP5/DPY30	51021	50 μg
MLL3/WDR5/Ash2L/RbBP5/DPY30	100323	50 μg
MLL4/WDR5/Ash2L/RbBP5/DPY30	100324	50 μg
MLL1/WDR5/Ash2L/RbBP5	51022	50 μg
MLL1 Complex Chemiluminescent Assay Kit	53008	96 rxns.

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** or Fax **1.858.481.8694**

Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com



6042 Cornerstone Court W, Ste B
 San Diego, CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

TROUBLESHOOTING GUIDE

Problem	Possible Cause	Solution
Luminescence signal of positive control reaction is weak	MLL4 enzyme has lost activity	Enzyme loses activity upon repeated freeze/thaw cycles. Use fresh enzyme (MLL4 complex, BPS Bioscience #100324). Store enzyme in single-use aliquots. Increase time of enzyme incubation. Increase enzyme concentration.
	Antibody reaction is insufficient	Increase time for antibody incubation. Avoid freeze/thaw cycles of antibodies.
	Incorrect settings on instruments	Refer to instrument instructions for settings to increase sensitivity of light detection. See section on "Reading Chemiluminescence" above.
	Chemiluminescent reagents mixed too soon	Chemiluminescent solution should be used within 15 minutes of mixing. Ensure both reagents are properly mixed.
Luminescent signal is erratic or varies widely among wells	Inaccurate pipetting/technique	Run duplicates of all reactions. Use a multichannel pipettor. Use master mixes to minimize errors.
	Bubbles in wells	Pipette slowly to avoid bubble formation. Tap plate lightly to disperse bubbles; be careful not to splash between wells.
Background (signal to noise ratio) is high	Insufficient washes	Be sure to include blocking steps after wash steps. Increase number of washes. Increase wash volume. Increase Tween-20 concentration to 0.1% in TBST.
	Sample solvent is inhibiting the enzyme	Run negative control assay including solvent. Maintain DMSO level at <1% Increase time of enzyme incubation.
	Results are outside the linear range of the assay	Use different concentrations of enzyme (MLL4 complex, BPS Bioscience #100324) to create a standard curve.

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401** or Fax **1.858.481.8694**
 Or you can Email us at: info@bpsbioscience.com
 Please visit our website at: www.bpsbioscience.com