



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Data Sheet

Fluorogenic MMP10 Assay Kit

Catalog #79986
Size: 96 reactions

BACKGROUND: MMP10 (matrix metalloproteinase 10), also known as Stromelysin-2, is a member of the matrix metalloproteinase (MMP) family involved in the degradation of the extracellular matrix. MMP10 moderates inflammation by controlling macrophage activation and is linked to several lung disorders, including pulmonary fibrosis and cystic fibrosis. MMP10 has also been shown to cleave the mutant huntingtin protein associated with Huntington's disease.

DESCRIPTION: The *Fluorogenic MMP10 Assay Kit* is designed to measure MMP10 activity for screening and profiling applications, in a homogeneous assay with no time-consuming washing steps. The kit comes in a convenient 96-well format, with purified MMP10 enzyme, fluorogenic substrate, and MMP assay buffer for 100 enzyme reactions.

COMPONENTS:

Catalog #	Component	Amount	Storage	
100565	MMP10, His-tag	7 µg	-80°C	Avoid freeze/ thaw cycles!
79919	1 mM MMP Substrate	10 µl	-80°C	
79917	1X MMP Assay Buffer 1	25 ml	-20°C	
79685	Black, low binding black microtiter plate	1	Room Temperature	

APPLICATIONS: Great for studying enzyme kinetics and HTS applications.

STABILITY: At least one year from date of receipt when stored as directed.

REFERENCE(S):

- Batra, J. *et al.* Matrix Metalloproteinase-10/TIMP-2 Structure and Analyses Define Conserved Core Interactions and Diverse Exosite Interactions in MMP/TIMP Complexes. *PLoS One*. 2013 Sep 20; **8(9):e75836**.
- McMahan, R.S., *et al.* Stromelysin-2 (MMP10) Moderates Inflammation by Controlling Macrophage Activation. *J. Immunol.* August 1, 2016, **197(3): 899-909**.

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401**, Fax **1.858.481.8694**

Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com

MATERIALS OR INSTRUMENTS REQUIRED BUT NOT SUPPLIED:

Fluorescent microplate reader capable of reading $\lambda_{exc}/\lambda_{em}=328\text{ nm}/393\text{ nm}$

ASSAY PROTOCOL:

All samples and controls should be tested in duplicate.

- 1) Dilute 1 mM MMP substrate 1:100 in 1X assay buffer, to make a 10 μM solution. Dilute only enough as is required for the assay. Store remaining 1 mM substrate in aliquots at -80°C .
- 2) Prepare the substrate solution: N wells \times (20 μl 1X assay buffer + 5 μl diluted (10 μM) MMP Substrate).
- 3) Add 25 μl of the substrate solution to each well (Final concentration of the MMP substrate in a 50 μl reaction is 1 μM).

Component	Positive Control	Test Sample	Blank
Substrate solution	25 μl	25 μl	25 μl
Test Inhibitor	–	5 μl	–
Inhibitor buffer (usually 10% DMSO in 1X assay buffer)	5 μl	–	5 μl
MMP10 (3.2 ng/ μl)	20 μl	20 μl	–
1X Assay Buffer	–	–	20 μl
Total	50 μl	50 μl	50 μl

- 4) Prepare the inhibitor solution.

The final concentration of DMSO in the assay should not exceed 1%. If the inhibitor compound is dissolved in DMSO, make a 100-fold higher concentration of the compound than the highest concentration you want to test in DMSO. Then make a 10-fold dilution in 1X assay buffer (at this step the compound concentration is 10-fold higher than the final concentration in 10% DMSO). To determine an IC_{50} or to test lower concentrations of the compound, prepare a series of further dilutions in 1X assay buffer containing 10% DMSO (the final concentration of the DMSO will be 1% in all samples).

If the inhibitor compound is dissolved in water, make a solution of the compound 10-fold higher than the final concentration in 1X assay buffer (with 1 mM DTT).

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.


To place your order, please contact us by Phone **1.858.202.1401**, Fax **1.858.481.8694**

Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com

- 5) Add 5 μ l inhibitor solution to each well designated "Test Sample." Add 5 μ l of inhibitor buffer (the same solution as the test sample, but without inhibitor) to "Blank" and "Positive Control" wells.
- 6) Thaw MMP10 on ice. Upon first thaw, briefly spin tube containing enzyme to recover the full content of the tube. Aliquot MMP10 into single use aliquots. Store remaining undiluted enzyme in aliquots at -80°C . Note: MMP10 enzyme is sensitive to freeze/thaw cycles. Do not re-use diluted enzyme.
- 7) Dilute MMP10 in 1x assay buffer at 3.2 ng/ μ l (64 ng per reaction).
- 8) Add 20 μ l diluted MMP10 enzyme solution to wells designated as "Positive Control" and "Test Sample." Add 20 μ l 1X assay buffer to the "Blank" wells.
- 9) Incubate at room temperature for 30 minutes. Measure the fluorescence intensity in a microtiter plate-reading fluorimeter capable of excitation at a wavelength 328 nm and detection of emission at a wavelength 393 nm. The fluorescence intensity can also be measured kinetically. "Blank" value is subtracted from all other values.

EXAMPLE OF ASSAY RESULTS:



Inhibition of MMP10 enzyme activity by NNGH, measured using the *Fluorogenic MMP10 Assay Kit (BPS Bioscience #79986)*. Fluorescence intensity was measured using a Tecan fluorescent microplate reader. *Data shown is lot-specific. For lot-specific information, please contact BPS Bioscience, Inc. at info@bpsbioscience.com*

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401**, Fax **1.858.481.8694**

Or you can Email us at: info@bpsbioscience.com

Please visit our website at: www.bpsbioscience.com



6042 Cornerstone Court West, Ste. B
San Diego CA 92121
Tel: 1.858.202.1401
Fax: 1.858.481.8694
Email: info@bpsbioscience.com

RELATED PRODUCTS

<u>Product</u>	<u>Cat. #</u>	<u>Size</u>
MMP1, His-Tag (Human)	80214	20 µg
MMP2, His-Tag (Human)	80213	20 µg
MMP3(K45E), His-Tag (Human)	11346	100 µg
MMP8, His-Tag (Human)	100552	100 µg
MMP9(Q279R), His-Tag (Human)	80215	20 µg
Fluorogenic MMP3 (K45E) Assay Kit	79907	384 rxns.
Fluorogenic MMP2 Assay Kit	79918	96 rxns.
Fluorogenic MMP8 Assay Kit	79929	96 rxns.
Fluorogenic MMP9 (Q279R) Assay Kit	79915	96 rxns.

OUR PRODUCTS ARE FOR RESEARCH USE ONLY. NOT FOR DIAGNOSTIC OR THERAPEUTIC USE.

To place your order, please contact us by Phone **1.858.202.1401**, Fax **1.858.481.8694**

Or you can Email us at: [**info@bpsbioscience.com**](mailto:info@bpsbioscience.com)

Please visit our website at: [**www.bpsbioscience.com**](http://www.bpsbioscience.com)