



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 



Q-Plex™ Mouse HS

Version 1.1
For Research Use Only
Not For Use in Diagnostic Procedures







Symbol	Explanation
	Catalog Number
	Lot Number
	Use By YYYY-MM-DD
	Temperature Limitation
	Manufacturer
	Keep Away from Sunlight

TABLE OF CONTENTS

Name and Intended Use	2
Principle of the Assay	3
Sample Collection and Storage	4
Warnings and Precautions	5
Kit Contents & Storage	6
Materials Required but not Supplied	7
Assay Preparation	8
Assay Procedure	9
Analyzing a Q-Plex™ Image	12
Performance Characteristics	13
Typical Data	13
Sensitivity	13
Precision	15
Dilutional Linearity	15
Specificity	16
Calibration	17
Appendix A: Plate Washing Methods	18
Automated Wash Method	18
Multichannel Pipette Wash Method	19
Abbreviated Protocol	20
Preparation	20
Running the Assay	20
Plate Diagram	21
Notes	22



Manufactured and Distributed by:

Quansys Biosciences

365 North 600 West Logan, Utah 84321, USA

TEL: (888) 782-6797

E-MAIL: info@quansysbio.com

For International Distributors see:

quansysbio.com/distributors

QUANSYS 
B I O S C I E N C E S

NAME AND INTENDED USE

Q-Plex™ Mouse HS

The Q-Plex Mouse HS is a chemiluminescent Enzyme-Linked Immunosorbent Assay (ELISA) intended for quantitative concurrent measurement of different proteins in serum or plasma samples. The kit accompanying this manual will contain a subset of the following analytes.

Mouse Analyte Targets

Eotaxin	IL-1 β	IL-10	MDC
GM-CSF	IL-2	IL-12p70	MIP-1 α
IFN α	IL-3	IL-13	RANTES
IFN β	IL-4	IL-17	TARC
IFN γ	IL-5	KC	TCA
IL-1 α	IL-6	MCP-1	TNF α

PRINCIPLE OF THE ASSAY

This multiplex assay is based on the microplate sandwich enzyme immunoassay technique. Samples or calibrators are pipetted into wells of a 96-well plate arrayed with immobilized analyte-specific antibodies that capture the protein of interest. After washing away any unbound protein, a mixture that contains biotinylated analyte-specific antibodies is added. The biotinylated antibodies complete the sandwich for each specific arrayed analyte. After washing away unbound biotinylated antibody, streptavidin-horseradish peroxidase (SHRP) is added. Following an additional wash, the amount of SHRP bound to each location of the array is proportional to the amount of protein initially captured.

The amount of conjugated enzyme on each location of the array is measured with the addition of a chemiluminescent substrate.

SAMPLE COLLECTION AND STORAGE

The sample collection and storage conditions are intended as general guidelines. The Clinical and Laboratory Standards Institute (CLSI GP44-A4) recommends the following storage conditions for samples: Samples should be stored at room temperature no longer than 8 hours. If the assay will not be completed within 8 hours, the samples should be refrigerated at +2°C to +8°C. If the assay will not be completed within 48 hours, or if the samples will be stored beyond 48 hours, samples should be frozen at -20°C or lower. Samples should not be repeatedly frozen and thawed. Frozen samples must be mixed well after thawing and prior to testing. Do not use bacterially contaminated samples. It is the responsibility of the individual laboratory to use all available references and/or its own studies to determine its own specific stability criteria.

Cell Culture Supernates – Remove particulates by centrifugation and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Serum – Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Plasma – Collect plasma using EDTA, Citrate, or Heparin as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

WARNINGS AND PRECAUTIONS

1. Read all instructions before beginning test.
2. For research use only. Not for use in diagnostic procedures.
3. The kit should not be used beyond the expiration date on the kit label.
4. If running multiple kits, a calibration curve must be included for each 96-well plate. The curve from one plate cannot be used to calculate sample values from other plates.
5. Do not mix or substitute reagents with those from other kits or lots.
6. Pre-wet pipette tips 3 times by drawing up the liquid into the pipette and then dispensing back into the original vessel prior to the addition of samples, calibrators, or external controls to the microplate.
7. Load all calibrators, controls, and samples into the microplate within 10 minutes of each other.
8. Be exact with incubation times, particularly the streptavidin-horseradish peroxidase (SHRP) incubation.
9. Be exact when mixing Substrate A and B+ and mix thoroughly.
10. Do not allow the substrate or SHRP to be exposed to UV light, as this may degrade these reagents.
11. Do not allow the plate to dry out between steps.
12. All products are carefully validated; however due to the variability encountered in biological sample matrices, the possibility of interference or sample matrix effects cannot be excluded.

KIT CONTENTS & STORAGE

Unopened Kit - Store at 2-8°C. Do not use past kit expiration date. Do not re-use.

Part/Description	Storage of opened/reconstituted material
Q-Plex™ Array Microplate 1 arrayed and blocked 96-well polystyrene micro-titer plate	↓ 2-8°C until kit expiration
Wash Buffer Concentrate (20X) Liquid, 50 mL/vial of a concentrated solution of buffered surfactant	
Sample Diluent Liquid, 10 mL/vial of a buffered protein solution	
Detection Mix Lyophilized, 6 mL/vial of biotinylated antibodies in a buffered protein solution with preservatives	↓ 2-8°C until kit expiration Discard unused reconstituted calibrator.
Calibrator Lyophilized, recombinant antigens in a buffered protein base	
Streptavidin-HRP Liquid, 6 mL/vial of streptavidin-conjugated horseradish peroxidase	⚠ Do not expose to UV light. ↓ 2-8°C until kit expiration
Substrate A Liquid, 3.5 mL/vial	⚠ Do not expose to UV light. ↓ Store mixed substrate solution at room temperature (20-25°C) for up to 1 week. ↓ Store unmixed solutions at 2-8°C until kit expiration.
Substrate B+ Liquid, 3.5 mL/vial	
Plate Seals (3) Adhesive strips	Non-perishable

MATERIALS REQUIRED BUT NOT SUPPLIED

In addition to the kit contents listed, the following materials are required to run this assay:

1. Multichannel pipette (20-200 μL) and/or single channel pipettes (20-1000 μL) with appropriate tips
2. 10 mL serological pipette
3. Polypropylene tubes or polypropylene 96-well plate(s) for sample and calibrator preparation
 - a. Example: Nunc® MicroWell™ 96-Well Plates, Polypropylene, 249944; Eppendorf® LoBind Protein or Genomic Microcentrifuge Tubes, 022431102
4. Q-View™ Imager and Q-View™ Software
5. Microplate shaker capable of 500 RPM
6. Deionized water
7. Microplate washer (An optional multichannel pipette protocol is provided in Appendix A)
8. 1 liter graduated cylinder for the preparation of wash buffer.


ASSAY PREPARATION

1. Install Q-View Software on any computers that will be used for analysis or operating a Q-View Imager.
2. Set up the imager. For imager-specific instructions, see www.quansysbio.com/support.
3. Set the microplate shaker to 500 RPM.
4. Allow all kit components to come to room temperature (20-25°C) prior to use.
5. Prepare Wash Buffer: Place 50 mL of Wash Buffer Concentrate (20X) into 950 mL deionized water and mix thoroughly.
6. Prepare Calibrator: Reconstitute lyophilized Calibrator using Sample Diluent with the volume specified on the In-Kit Certificate of Analysis accompanying the kit. Allow to sit for 5 minutes. Mix thoroughly. Use prepared Calibrator within 30 minutes.
7. Prepare Detection Mix: Reconstitute lyophilized Detection using 6 mL of DI water. Allow to sit for 5 minutes. Mix thoroughly.
8. Prepare chemiluminescent substrate: Combine 3 mL of Substrate A with 3 mL of Substrate B+ and mix gently. **Do not expose to UV light. Store at room temperature (20-25°C) after mixing.**

ASSAY PROCEDURE

Allow all reagents to equilibrate to room temperature (20-25°C) before use and prepare as directed in the previous sections. It is recommended that all calibrators, samples, and controls be assayed in duplicate.

- Using the previously prepared Calibrator (see “Assay Preparation” step 6), prepare an 8-point calibration curve (7 points plus 1 blank) in either polypropylene tubes or a polypropylene 96-well plate (see “Materials Required but Not Supplied” for example tubes and plates).
 - Pipette 200 μL of prepared Calibrator into the first tube or well.
 - Pipette 120 μL of Sample Diluent into the other 7 tubes or wells.
 - Transfer 60 μL of undiluted prepared Calibrator from the first tube or well into the second, mix thoroughly, and repeat the transfer from tube to tube or well to well for 5 more points, leaving the last tube or well without any prepared Calibrator. This process is diagrammed below. The undiluted prepared Calibrator serves as the high point of the calibration curve. The Sample Diluent serves as the blank.



- Prepare samples by diluting 1:2 (1 part sample to 1 part Sample Diluent) in either polypropylene tubes or a polypropylene 96-well plate in a volume sufficient to provide 50 μL of diluted sample per well.

Note: If you anticipate that your analyte concentrations will be greater than the high point of the calibration curves as reported on the In-Kit Certificate of Analysis, use Sample Diluent to dilute your samples further.

3. Add 50 μL per well of the prepared calibration curve, controls, and samples to the Q-Plex™ Array Microplate. Load all samples and calibration curve to the plate within 10 minutes.
4. Cover the plate with a plate seal and incubate on a plate shaker set to 500 RPM for 180 minutes at room temperature (20-25°C).
5. Wash the plate 3 times (see Appendix A).
6. Add 50 μL per well of Detection Mix, cover with a new plate seal, and return to the plate shaker set to 500 RPM for 120 minutes at room temperature (20-25°C).
7. Wash the plate 3 times (see Appendix A).
8. Add 50 μL per well of Streptavidin-HRP, cover with a new plate seal, and return to the plate shaker set to 500 RPM for 20 minutes at room temperature (20-25°C).
9. Wash the plate 6 times (see Appendix A).
10. Add 50 μL per well of previously prepared substrate (see “Assay Preparation” step 7). Image plate immediately. Wait no longer than 5 minutes to commence imaging.

Note: If imaging cannot commence immediately, protect the plate from drying for up to 15 minutes by dispensing 100 μL of wash buffer into each well of the plate prior to adding the mixed substrate. When ready to image, remove the wash buffer from the plate and add the substrate.

11. Place the plate in the Q-View Imager.
12. Open Q-View Software, create or open a project, and click Acquire Image.
13. For imager-specific recommended exposure times, see www.quansysbio.com/support.
14. Click the Capture Image(s) button. Users may perform Well Assignment while images are being captured.

Note: Details about these imaging steps are available in the Q-View Software Manual viewable at www.quansysbio.com/support or within Q-View Software under **Support > Manual**.

15. Dispose of all used and unused materials in accordance with local regulations.

ANALYZING A Q-PLEX™ IMAGE


The following summarizes a general workflow for analyzing a Q-Plex image in Q-View Software. Each of these steps is described in greater detail in the Q-View Software and Imager manuals, viewable at www.quansysbio.com/support or within Q-View Software under **Support > Manual**.

1. Acquire or import an image into Q-View Software as previously described.
2. Enter the **Software Product Code** (found on the In-Kit Certificate of Analysis) into the **Product Code** field of the software.
3. **Image Processing**: Align the plate overlay as follows:
 - a. Set the overlay: If using the Auto-Set Plate Overlay feature, this will occur automatically. Otherwise, go to **Overlay Options > Set Plate Overlay**
 - b. To visualize bright or dim spots, optimize the display using **Image Options > Adjust Gamma** (does not affect the data).
 - c. Optimize overlay alignment: Go to **Overlay Options > Adjust Plate Overlay** to pivot the overlay, **Adjust Well** and **Adjust Spot** to move individual wells and spots, then **Auto-Adjust Spots** to automatically snap each circle of the overlay to the nearest spot of the image beneath.
4. **Well Assignment**: Label wells as samples, calibrators, controls, or negatives (blank), and specify their dilution factors. Use Templates to quickly assign layouts that are repeated often or export the layout as a .csv file.
5. **Data Analysis**: Once **Image Processing** and **Well Assignment** are complete, select **Data Analysis**. Click **Perform Analysis** to generate charts, concentrations, and statistics based on optimal regression settings provided by the product definition.

PERFORMANCE CHARACTERISTICS

Typical Data

Representative data from one lot of kits are shown below to demonstrate typical calibrator curve performance.



Sensitivity

The upper limit of quantitation (ULOQ) is the highest concentration at which the recovery of the analyte is within 80-120% of the known value.

The lower limit of quantitation (LLOQ) is the lowest concentration at which the recovery of the analyte is within 75-125% of the known value and the CV of the calculated concentration is <25%.

The lower limit of detection (LLD) is the concentration that is two standard deviations above the average signal of twenty blank (zero) replicates.

Analyte	ULOQ (pg/mL)	LLOQ (pg/mL)	LLD (pg/mL)
Eotaxin	2392.7	3.53	1.98
GM-CSF	467.7	0.61	0.43
IFN α	22400.0	57.1	28.55
IFN β	736.8	1.06	0.42
IFN γ	414.6	0.54	0.12
IL-1 α	540.6	0.72	0.068
IL-1 β	2052.3	2.51	0.16
IL-2	280.3	0.38	0.053
IL-3	58.0	0.08	0.022
IL-4	329.9	0.43	0.050
IL-5	419.3	0.54	0.054
IL-6	522.0	0.68	0.18
IL-10	533.8	0.70	0.090
IL-12p70	508.5	0.65	0.10
IL-13	402.4	0.53	0.076
IL-17	613.7	0.77	0.16
KC	403.7	0.54	0.098
MCP-1	8740.4	11.12	4.25
MDC	459.0	0.61	0.21
MIP-1 α	850.2	1.16	0.50
RANTES	1893.7	2.46	0.85
TARC	15284.6	19.88	1.74
TCA-3	1137.5	1.60	0.56
TNF α	533.2	0.67	0.15

Precision

Intra-Assay Precision – 3 samples containing and/or spiked with various concentrations of each analyte were tested 20 times on 1 plate to assess intra-assay precision. The average %CV of all samples is reported.

Inter-Assay Precision – 3 samples containing and/or spiked with various concentrations of each analyte were tested on at least 10 individual assays by at least 3 users. The average %CV of all samples is reported.

Dilutional Linearity

Serum and plasma samples containing and/or spiked with each analyte were diluted with sample diluent at dilutions of 1:2, 1:4, 1:8, 1:16 and assayed. The highest, lowest, and average dilution linearity are reported.

Analyte	Average Intra-Assay Precision (%CV)	Average Inter-Assay Precision (%CV)	Linearity Range	Linearity Average
Eotaxin	5.7%	3.6%	77-109%	84%
GM-CSF	7.5%	4.9%	72-115%	94%
IFN α	7.1%	8.7%	95-117%	109%
IFN β	8.1%	4.5%	99-122%	110%
IFN γ	5.5%	3.3%	83-133%	110%
IL-1 α	7.4%	3.3%	80-130%	108%
IL-1 β	9.3%	3.1%	74-108%	81%
IL-2	6.6%	5.1%	76-103%	88%
IL-3	7.0%	4.7%	89-130%	106%
IL-4	6.6%	6.2%	75-132%	98%
IL-5	5.2%	6.6%	75-122%	103%
IL-6	6.1%	7.3%	89-124%	106%
IL-10	3.0%	7.5%	86-130%	108%
IL-12p70	10.8%	5.5%	88-129%	107%
IL-13	13.9%	13.8%	80-133%	103%

Analyte	Average Intra-Assay Precision (%CV)	Average Inter-Assay Precision (%CV)	Linearity Range	Linearity Average
IL-17	6.1%	6.5%	74-101%	92%
KC	5.2%	4.3%	98-185%*	135%
MCP-1	9.2%	5.5%	76-129%	96%
MDC	4.7%	7.4%	88-107%	99%
MIP-1 α	6.7%	10.9%	71-125%	86%
RANTES	5.5%	5.0%	90-123%	108%
TARC	8.5%	13.5%	71-127%	108%
TCA-3	8.1%	13.3%	96-134%	108%
TNF α	5.2%	6.9%	87-113%	104%

*KC does not dilute well from a 1:2 to 1:4 dilution. From a 1:4-1:16, the range of linearity seen is 98%-132%. Samples tested at a 1:2 on KC should not be compared to samples tested at other dilutions.

Specificity

Specificity of the array was assessed by testing each recombinant mouse antigen at the concentration of the highest calibrator in the multiplexed assay. The cross-reactivity of all antigens on each assay was less than 1%.

Calibration

This immunoassay is calibrated against highly purified *E. coli*-expressed recombinant mouse antigens prepared as a reference calibrator at Quansys Biosciences.

The Q-Plex reference calibrator was evaluated against the NIBSC/WHO International Standards, intended as bioassay standards. The NIBSC/WHO standards are assigned arbitrary unitage in U/mL. The ratios of International Units of biological activity per mL (IU/mL) of NIBSC standard relative to pg/mL of Q-Plex calibrator are shown in the table below. To convert Q-Plex concentrations (pg/mL) to biological activity relative to the WHO International Standard, multiply the Q-Plex concentration by the ratio provided.

Antigens	NIBSC/WHO Catalog Number	Concentration Ratio
IL-1 α	93/672	4.55
IL-1 β	93/668	0.95
IL-2	93/566	0.20
IL-3	91/662	0.20
IL-4	91/656	0.0024
IL-6	93/730	6.87
TNF α	88/532	0.85
GM-CSF	91/658	0.11

APPENDIX A: PLATE WASHING METHODS

Automated Wash Method

1. Use a program that will aspirate and dispense 300-400 μL wash buffer.



2. Ensure that the plate washer has a program that will not scratch the bottom of the microplate but will prevent plate drying. This is critical to prevent damage to the capture antibody arrays. The simplest method to avoid plate drying is to leave a small, uniform amount (1-3 μL) of wash in the well after the final aspiration and add the next reagent to the plate as quickly as possible.
3. Leaving a uniform amount of wash in the wells can be accomplished with an automated washer by slightly increasing the aspiration height of the washer head.

For example:

Process	Distance	Steps on a Biotek ELX-405
Aspiration Height	3.81 mm	30
Aspiration Position	1.28 mm from center	-28
Dispense Height	15.24 mm	120

No soak or shake cycles are needed

4. Connect the prepared wash buffer to your automatic plate washer.
5. Run 1-2 priming cycles to ensure that wash buffer is running through the plate washer. When the buffer has run through the machine, the waste will be foamy.
6. In a spare microtiter plate, dispense 100 μL wash buffer, ensure that all pins dispense uniformly, then aspirate and ensure that all pins aspirate uniformly. This will ensure that all pins are functioning.
7. Prime the plate washer 1 time before the first wash step. When running the assay, perform the wash 3 or 6 times according to the protocol.

Multichannel Pipette Wash Method

1. Just prior to washing, pour the prepared wash buffer into a trough or tray.
2. After each incubation, flick the solution out of the plate over a waste container before starting the wash protocol.
3. Using a multichannel pipette, dispense 400 μL of wash buffer into each of the wells used in the test.
4. Aggressively flick the wash buffer out over a waste container.
5. This washes the plate 1 time. When the assay procedure calls for 3 or 6 washes, repeat steps 3-4 accordingly.
6. Tap the plate upside down on a paper towel to remove any residual wash.
7. Proceed immediately to dispense the next solution so drying does not occur.

ABBREVIATED PROTOCOL

Preparation


1. Install Q-View Software (page 8).
2. Set up the imager (page 8).
3. Set up microplate washer (page 18) and shaker (page 8).
4. Reconstitute and prepare reagents (page 8).

Running the Assay

5. Prepare the calibration curve using the prepared Calibrator and Sample Diluent according to the In-Kit Certificate of Analysis (page 9).
6. Prepare the samples with Sample Diluent (page 9).
7. Load the calibration curve and samples into the plate. Shake for 180 minutes (500 RPM) at room temperature (page 10).
8. Wash the plate 3 times, add the Detection Mix, and shake for 120 minutes (500 RPM) at room temperature (page 10).
9. Wash the plate 3 times, add the Streptavidin-HRP, and shake for 20 minutes (500 RPM) at room temperature (page 10).
10. Wash the plate 6 times and add the prepared Substrate (page 10).
11. Capture and analyze image of the plate (page 11).

We take great care to ensure that customers have success using our products and services. If you have further questions about running the assay, data analysis, troubleshooting, or our products or services, please contact us at 1-888-QUANSYS (1-888-782-6797) or at support@quansysbio.com.

PLATE DIAGRAM



Q-Plex™ ARRAY

For technical support, email: support@quansysbio.com

112417MS

365 North 600 West, Logan, Utah 84321
Tel: 1-888-782-6797 | Fax: (435)750-6869 | www.quansysbio.com