



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 



Place your order with CEDARLANE® or your local distributor.

Please contact CEDARLANE® for lot specific information.

**PE-Cy5 Rat IgG₁
Isotypic Control**

CLCR106

LOT:

DESCRIPTION: PE-Cy5 Rat IgG₁ isotypic control.

SIZE: 0.5 ml

ANTIBODY CONTENT: 50 µg

APPLICATION :

This isotype control is tested by flow cytometry using freshly harvested mouse splenocytes and/or human peripheral blood lymphocytes (PBL). When using isotype controls as negative controls, it is recommended that the amount of isotype control and testing antibody used be equivalent.

BUFFERS & PRESERVATIVE:

This reagent is provided in phosphate buffered saline and contains 0.1% sodium azide (NaN₃) as a preservative. A highly purified grade of BSA has been added as a stabilizing protein to bring the final protein concentration to 4-5 mg/ml after conjugation.

Continued Overleaf...

For more information or to place an order please contact...

CEDARLANE®
LABORATORIES LIMITED



toll free: 1-800-268-5058
in North America

phone: (905) 878-8891 • fax: (905) 878-7800

5516 - 8th Line, R.R.#2, Hornby, Ontario, CANADA L0P 1E0

or visit our website for a list of our international distributors including contact information

website: www.cedarlanelabs.com • e-mail: info@cedarlanelabs.com

STORAGE/STABILITY:

Store reagents at 2-8°C. Light exposure should be avoided when using fluorochrome-conjugated antibodies. Use dim light during handling, incubation with cells and prior to analysis. It is recommended that cells be analyzed within 18 hours of staining. If the reagent is being diluted, it is recommended that only the quantity to be used within one week be diluted. Check label for expiry date.

FOR RESEARCH USE ONLY

TN 11/8/04