



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

TECHNICALLY *Speaking*

Place your order with CEDARLANE® or your local distributor.

Please contact CEDARLANE® for lot specific information.

PE-Cy7 Rat IgG_{2a} Isotypic Control

CLCR2A12

LOT: 01021102

DESCRIPTION: PE-Cy7 Rat IgG_{2a} isotypic control

SIZE: 0.5 ml

ANTIBODY CONTENT: 50 µg

CLONE: LO-DNP-16

APPLICATION: This isotype control is provided in liquid form for use in flow cytometry. 1 µg is the maximum amount of reagent required for 1×10^6 cells. Lesser amounts of reagent may be sufficient and it is recommended that the customer determine the optimum amount for each application. When using isotype controls as negative controls, it is recommended that the amount of isotype control and the testing antibody used be equivalent.

BUFFERS & PRESERVATIVE: This reagent is provided in phosphate buffered saline and contains 0.1% sodium azide (NaN_3). A highly purified grade of BSA has been added as a stabilizing protein to bring the final concentration to 4-5 mg/ml after conjugation.

Continued Overleaf...

For more information or to place an order please contact...

CEDARLANE®
LABORATORIES LIMITED



toll free: 1-800-268-5058
in North America

phone: (905) 878-8891 • fax: (905) 878-7800

5516 - 8th Line, R.R.#2, Hornby, Ontario, CANADA L0P 1E0

or visit our website for a list of our international distributors including contact information
website: www.cedarlanelabs.com • e-mail: info@cedarlanelabs.com

Using formalin to fix cells following immunofluorescent staining may cause the degradation of tandem fluorochromes. Cells stained with PE-Cy7 should be analyzed by flow cytometry within 18 hours following fixation.

STORAGE/STABILITY: Stable at 4°C. **DO NOT FREEZE.** Avoid prolonged exposure to light. If the reagent is being diluted, it is recommended that only the quantity to be used within one week be diluted. Check label for expiry date.

FOR RESEARCH USE ONLY

CH 04/18/05