



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Axygen Inc.**

Registered  
ISO 9001:2008  
ISO 13485-2003

## Product Description

**Catalog Number:** MCT-175-C-S

**Product Description:** Axygen® Microcentrifuge Tube, PP, Sterile, 1.7 mL

**Component Materials:**

Tube: Polypropylene. Color: Clear.

**Attributes:**

Tube Overall Length with Cap Closed: 20 mm

Tube OD: 7 or 8 mm

**Autoclaving:**

This product is single use and provided sterile therefore routine autoclaving is not expected. The product can be autoclaved at 121°C and 15 psi for 15 minutes.

**Heating / Cooling:**

This product can be heated to 130°C. This product can be cooled to -20°C.

**Vacuum / Pressure:**

This product is not rated for vacuum or pressure.

**Sterilization:**

This product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products - Requirements for validation and routine control - Radiation sterilization.*

Sterility Assurance Level (SAL): 10<sup>-6</sup>

**Pyrogens:**

This product is tested according to ANSI/AAMI ST 72:2002 (R)2010 *Bacterial endotoxin test methodologies, routine monitoring and alternatives to batch testing.*

Results: < 0.1 EU/mL

**RNase/DNase Testing:**

This product is tested for absence of detectable RNase/DNase contamination.

**TSE/BSE:**

This product complies with the latest revision of EMEA/410/01, "Note for guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human veterinary medicine products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMEA/410/01.

O-rings are made from ingredients that have been prepared solely from synthetic and/or plant sources. It does not contain any materials of animal origin however no warranty is expressed or implied.

**Lot Number Designation:**

The product lot number shall be 6 digits followed by a "-" and 3 digits. The first two digits of the lot number shall be the last two digits of the current year. The next two digits shall represent the current month. The last two of the six digits shall represent the current date of the month. The last three numbers are preceded by a "-" and represents the actual mold number.

**Expiration Statement:**

Axygen's practice is to recommend product use within three years from the date of manufacture.

Rev No: 2