



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

## Product Description

**Catalog Number:** T-350-C

**Product Description:** T-350 Test Cert 5-23-2012

**Component Materials:**

- Tip - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Rack - Polypropylene with orange colorant, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets CONEG req.) color concentrate.

**Product Dimensions:**

Overall tip length	-	1.300"	I.D. small orifice	-	.015"
Tolerances	-	+/- .050"			

**Special Feature:**

The entire unit is autoclavable.

**RNase/DNase Testing:**

This product has been tested and is free of any detectable RNase/DNase contamination.

**Pyrogens:**

The product has been tested and has met the criteria established in the *Guideline on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987)*. Results: less than 0.1 EU/mL

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification. The entire lot is sterilized as one unit.

Rev No: 1