



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 354484

Product Description: Corning® Biocoat™ 25cm² canted neck flask, collagen I coated, with filter vent cap

Component Materials:

| | | |
|------------|---|--|
| Collagen I | - | Rat Tail Tendon |
| Flask | - | Polystyrene |
| Cap | - | Polyethylene with color concentrate |
| Filter | - | 0.2µm Versapore acrylic copolymer membrane |

Bioburden:

This lot has been tested and found negative for Bacteria and Fungi.

Cell Attachment and Growth Characteristics:

This lot has been tested for the attribute of Cell Attachment and Growth utilizing human HT1080 cells, an attachment-dependent mammalian cell line in a serum free media.

Collagen I:

The dishes are coated with Collagen I which is a rat tail collagen, coated via a proprietary method.

Storage Conditions:

For optimal performance, the unopened product should be stored in a dry place at 4-30 degrees C (Do not freeze).

Handling:

The product is ready to use and does not require washing of the coated surface prior to use.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

| | |
|-------------------------|---|
| Visual Attributes: | Visual and microscopic examination of the product. |
| Bacteria & Fungi Free : | Turbidity test |
| Packaging: | Inspection for seal and barrier integrity, accurate labeling and correct product configuration. |

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian Date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 1