



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

**Registered**  
ISO 9001:2008

**Product Description**

**Catalog Number:** 11440

**Product Description:** Corning® 1 Liter Disposable Erlenmeyer flask, with accessories and diptube

**Component Materials:**

- Flask - Virgin Polycarbonate, meets *USP, Class VI* requirements for plastic containers and closures.
- Cap - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate.

**Accessories:**

- Tubing - C-Flex, Thermoplastic Elastomer, meets *USP, Class VI* requirements for plastic containers and closures.
- Connector - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Filter - .2µm Acrylic/Versapor meets *USP, Class VI* requirements for Biological Plastics.
- Diptube - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Reducer - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

**Product Dimensions:**

- |                                  |   |          |                         |   |              |
|----------------------------------|---|----------|-------------------------|---|--------------|
| Height of flask with cap         | - | 9.13 in. | Height of flask w/o cap | - | 8.08 in.     |
| Diameter of flask @ widest point | - | 5.03 in. | Tolerances              | - | +/- 0.05 in. |
| Diameter of cap                  | - | 1.91in.  |                         |   |              |

**Sterilization:**

The product has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*  
Sterility Assurance Level: SAL 10<sup>-6</sup>

**Pyrogens:**

The Erlenmeyer Flask has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing.*

Results: ≤ 0.1 EU/mL ( ≤ 4EU/device)

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product is manufactured and packaged with animal free materials.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Integrity Testing: Pull Force test of the accessory to 10lbs.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 4