



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

## Product Description

**Catalog Number:** 11705

**Product Description:** Corning® 50 mL centrifuge tube, self-standing, polypropylene, with accessories and diptube

**Component Materials:**

Centrifuge Tube - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

Cap - Virgin High Density Polyethylene, meets *USP Class VI* requirements for plastic containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate.

**Accessories:**

Tubing - C-Flex, Thermoplastic Elastomer, meets *USP, Class VI* requirements for plastic containers and closures.

Connector - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

Filter - 25mm/.2µm Acrylic/Versapor meets *Class VI* requirements for Biological Test for Plastics.

Diptube - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

**Product Dimensions:**

Length of Tube - 4.52 in.

Outer Diameter of tube - 1.24 in.

Length of tube w/cap - 5.07 in.

Tolerances - +/- 0.05 in.

**Sterilization:**

This lot has been irradiated and dosimetrically released based on ANSI/AAMI.ISO 11137 (AAMI TIR33) *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization.*

Sterility Assurance Level: SAL 10<sup>-6</sup>

**Pyrogens:**

The Centrifuge Tube has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72: *Bacterial Endotoxins - Test methodologies, routine testing, and alternative to batch testing.*

Results: ≤ 0.1 EU/mL (≤ 4EU/device)

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product is manufactured and packaged with animal free materials.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures

Integrity:

Pull Force Test of the accessory to 10lbs.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 3