



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3152

Product Description: Corning ® Spinner Flask, Disposable, 125ml

Component Materials:

- Vessel - Virgin Polystyrene, *meets USP, Class VI* requirements for plastic containers and closures, "complies with International Organization for Standardization (ISO) 10993-18, Biological Evaluation of Medical Devices - Part 18: Chemical Characterization of Materials (2005)."
- Caps - Polypropylene, *meets USP, Class VI* requirements for plastic containers and closures.
- Flange Base - Acrylonitrile Butadiene Styrene
- Magnet - Polystyrene encapsulated ALNICO magnet, 12800 Gauss induction

Product Dimensions:

- | | | | | | |
|-----------------------|---|-------------------|---------------------------|---|-------------------|
| Height of Vessel | - | 5.7 in. (14.47cm) | Top Opening of Vessel | - | 2.5 in. (6.35cm) |
| Diameter of Vessel | - | 2.5 in. (6.35cm) | Top Cap of Vessel | - | 70mm GL 70 |
| Footprint of Vessel | - | 4.4 in. (11.18cm) | Sidearm Opening of Vessel | - | .74 in (1.87cm) |
| Sidearm Cap of Vessel | - | 25mm GL 25 | Height of Paddle | - | 1.94 in. (4.92cm) |
| Width of Paddle | - | 1.56 in. (3.97cm) | Working Volume | - | 125ml minimum |
| Tolerances | - | +/- 0.05 | | | |

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR 33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.* Sterility Assurance Level: SAL 10⁻⁶

Pyrogens:

This vessel has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing.* results: ≤ 0.1 EU/mL (≤ 4EU/device)

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product is manufactured and packaged with animal free materials.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

- Visual Attributes: Visual examination of the product.
- Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 8