



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

# Product Description

**Catalog Number:** 3153

**Product Description:** Corning® Spinner Flask, Disposable, 500ml

**Component Materials:**

- Vessel - Virgin Polystyrene, *meets USP, Class VI* requirements for plastic containers and closures, "complies with International Organization for Standardization (ISO) 10993-18, Biological Evaluation of Medical Devices - Part 18: Chemical Characterization of Materials (2005)."
- Caps - Polypropylene, *meets USP, Class VI* requirements for plastic containers and closures.
- Flange Base - Acrylonitrile Butadiene Styrene
- Magnet - Polystyrene encapsulated ALNICO magnet, 12800 Gauss induction

**Product Dimensions:**

- |                       |                     |                           |                     |
|-----------------------|---------------------|---------------------------|---------------------|
| Height of Vessel      | - 8.0 in. (20.32cm) | Top Opening of Vessel     | - 3.6 in. (9.21cm)  |
| Diameter of Vessel    | - 3.6in. (9.14cm)   | Top Cap of Vessel         | - 100mm GL 100      |
| Footprint of Vessel   | - 5.5 in. (13.97cm) | Sidearm Opening of Vessel | - 1.5 in. (3.81cm)  |
| Sidearm Cap of Vessel | - 45mm GL 25        | Height of Paddle          | - 2.31 in. (5.87cm) |
| Width of Paddle       | - 1.94 in. (4.92cm) | Working Volume            | - 500ml minimum     |
| Tolerances            | - +/- 0.05          |                           |                     |

**Sterilization:**

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR 33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.* Sterility Assurance Level: SAL 10<sup>-6</sup>

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product is manufactured and packaged with animal free materials.

**Pyrogens:**

This vessel has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing.* Results: ≤ 0.1 EU/mL (≤ 4EU/device)

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

- Visual Attributes: Visual examination of the product.
- Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 8