



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated  
Life Sciences**

**Registered  
ISO 9001:2008**

## Product Description

**Catalog Number:** 3268

**Product Description:** Corning® CellSTACK® - 1 Chamber

**Component Materials:**

- Top Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Bottom Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.
- Cap - Virgin High Density Polyethylene, meets *USP, Class VI* requirements for plastic containers and closures. Heavy metal free (meets *CONEG* req.) color concentrate.
- Filter - .2µm/polyester non-woven membrane, meets *USP, Class VI* requirements for plastic containers and closures.
- Adhesive - Proprietary Acrylate, meets ISO-10993, Biocompatibility requirements and does not contain any animal products.

**Product Dimensions:**

- |                         |   |                  |                         |   |                     |
|-------------------------|---|------------------|-------------------------|---|---------------------|
| Overall Length          | - | 13.2 in. (335mm) | Overall Width           | - | 8.1 in. (206mm)     |
| Overall Height with cap | - | 2.0 in. (51mm)   | Tolerances              | - | +/- 0.1 in. (2.5mm) |
| Neck ID                 | - | 1.0 in. (26mm)   | Neck O.D. incl. Threads | - | 1.3 in. (32mm)      |
| Distance between plates | - | 0.67 in. (17mm)  |                         |   |                     |

**Total Cell Growth Area:**

636 cm<sup>2</sup>

**Recommended Working Volume:**

130 - 200 ml

**Sterilization:**

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR33) *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization.*

Sterility Assurance Level: SAL 10<sup>-6</sup>

**Pyrogens:**

The CellStack has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72:2002/(R)2010 *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternative to batch testing.* Results: ≤ 0.1 EU/mL (≤ 4EU/device)

**Surface Characterization:**

Surface is characterized to be hydrophilic and negatively charged. The negatively charged, carboxyl surface composition, has been optimized for cell attachment and growth.

**Cell Attachment and Growth Characteristics:**

The product has been tested for the attribute of cell attachment and growth utilizing an attachment- dependent mammalian cell line in a serum supplemented media.

**Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:**

This product does not contain materials of animal origin.

**Performance Testing:**

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Integrity Testing:

Forward pressurization of the product to 1.5 psi.

Cell Culture Treatment:

Wettability test using water to insure the presence of a hydrophilic surface.

Visual Attributes:

Visual examination of the product.

Packaging:

Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

**Lot Number Designation:**

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

**Corning Incorporated  
Life Sciences**

Registered  
ISO 9001:2008

# Product Description

**Catalog Number:** 3268

**Product Description:** Corning® CellSTACK® - 1 Chamber

**Serial Number Designation:**

12 Digit Serial Number: First 8 digits - Lot number (see lot number designation above); Next 3 digits CellSTACK serial number; Last digit - Stack designation.

Rev No: 17