



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3324

Product Description: Corning ® CellSTACK ® - Universal Cap, 1 solid overcap, 2 vent caps

Component Materials:

- Caps - Dow LDPE 993I Natural
- Air foil - Selig 1-19/S70A.
0.2µm GORE-TEX® PTFE membrane meets, *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

- | | | | | | |
|------------------------|---|--------------|------------------------|---|-----------|
| Diameter of solid cap | - | 0.690 in. | Diameter of vent cap | - | 0.770 in. |
| Height of solid cap | - | 0.562 in. | Height of vent cap | - | 0.561 in. |
| Sm diameter of airfoil | - | 0.250 in. | Lg diameter of airfoil | - | 0.750 in. |
| Tolerances | - | +/- 0.05 in. | Spacing of airfoil | - | 0.750 in. |

Sterilization:

The product is irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 (AAMI TIR 27) *Sterilization of health care products-Requirements for validation and routine control-Radiation sterilization.*

Sterility Assurance Level: SAL 10⁻⁶

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 2