



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3847

Product Description: Corning® 384-well, Fibronectin Coated, clear plate with lid

Component Materials:

Plate/Lid - Virgin Polystyrene, meets USP, Class VI requirements for plastic containers and closures.

Fibronectin - Human Plasma

Product Dimensions:

Length of Plate	-	5.030 in.	Width of Square Well @ Top	-	.143 in.
Width of Plate	-	3.365 in.	Width of Square Well @ Bottom	-	.105 in.
Depth of Well	-	.450 in.	Height with Lid	-	.637 in.
Tolerances of Dimensions	-	+/- .010 in.	Working Volume per well	-	20 µL to 75 µL

Bioburden:

This product has been tested for Bioburden contamination. Results: less than .52% of wells contaminated.

Pyrogens:

The product has been tested and has met the criteria established in the current version of ANSI/AAMI ST 72: 2002/ (R) 2010: *Bacterial Endotoxins - Test methodologies, routine monitoring, and alternatives to batch testing.* Results: ≤ 0.5 EU/mL (≤ 20EU/device)

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line in a serum supplemented media.

Fibronectin:

The plates are coated with fibronectin via a proprietary method, which creates a uniform, dense fibronectin network on the plastic surface. The fibronectin used for coating is purified from human donor plasma that has been tested negative for HBsAg and HIV-I, HIV-II and HCV antibodies.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for seal and barrier integrity, accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 2