



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Product Description

Catalog Number: 431730

Product Description: Corning® Storage Bottle, 60 mL, Octagonal, PET, with 18 mm HDPE Screw Cap, Sterile, 24/Tray, 96/Case

Component Materials:

- Bottle - PET, clear, non-cytotoxic, non-hemolytic, meets USP, Class VI requirements for plastic containers and closures.
- Cap - HDPE, natural, non-cytotoxic, non-hemolytic, meets USP, Class VI requirements for plastic containers and closures.

Attributes:

- Capacity - 60 mL
- Product Dimensions - 82 mm Height (with cap locked); 81 mm Height (w/o cap); 40 mm Width; 26 mm Cap OD; 18 mm Neck ID

Autoclaving:

This product is not suitable for autoclaving.

Heating / Cooling:

This product is designed for use at temperatures ranging from -40°C to 50°C.

Special Features:

This product has a tamper evident cap. The product is supplied with the cap in place but not locked. This product is manufactured in an environment maintained to the standards of a Class 100 clean room.

Sterilization:

This product is sterilized using an irradiation method compliant with the requirements of ANSI/AAMI/ISO 11137 to achieve a Sterility Assurance Level (SAL) of 10^{-6} .

Pyrogens:

This product is evaluated for pyrogens according to USP 31 and meets the requirements for classification as non-pyrogenic. Results less than 0.1 EU/mL.

Expiration Statement:

This product is defined to have an expiration date of 3 years from the date of manufacture.

Revision Date:

20 NOV 2013

Rev No: 1