



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Catalog Number: 4373

Product Description: Corning® 1536 well, Corning®CellBIND® surface, black plate, with clear flat bottom, high web, low base, without lid, with barcode labels

Component Materials:

Plate walls -Cyclic Olefin Co-Polymer, meets USP, Class VI requirements for plastic containers and closures. Black cyclic olefin co-polymer concentrate
Plate bottom -Cyclic Olefin Co-Polymer, meets USP, Class VI requirements for plastic containers and closures.
Barcode label -White polypropylene base, acrylic adhesive.

Special Feature

Barcode label (format 128) on A1 short side & H12 short side both centered.

Product Dimensions:

Length of Plate	-	5.030 in.	Width of Square Well @ Top	-	.070 in.
Width of Plate	-	3.365 in.	Width of Square Well @ Bottom	-	.059 in.
Depth of Well	-	.244 in.	Height W/O Lid	-	.315 in.
Tolerance of Dimensions	-	+/- .010 in.	Working Volume	-	10µL
Maximum well volume	-	15µL	Tolerance of Dimension For length	-	+.010 / -.015 in.

Surface Characterization – Surface is characterized to be hydrophilic and negatively charged, composed of ≥ 20% oxygen atoms. This surface composition has been developed for cell attachment and growth.

Optical Characteristics – This product is made of opaque black cyclic olefin co-polymer walls to minimize well to well crosstalk and background fluorescence and / or luminescence. The bottom is made of clear cyclic olefin co-polymer to permit direct microscopic viewing.

BSE/TSE – Product complies with the latest resin of EMA/410/01 “Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products” by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Sterility – Product has been sterilized and dosimetrically released per the requirements of ANSI/AAMI/ISO 11137, “Sterilization of health care products- Radiation”. Products meet a minimum Sterility Assurance Level (SAL) of 10⁻³.

Non-Pyrogenic – Tested and met the criteria established in the current version of ANSI/AAMI ST 72, ‘Bacterial Endotoxins – Test methodologies, routine monitoring, and alternatives to batch testing’. The acceptance level for product is ≤ 0.10 EU/ml or ≤ 4 EU/device.

Tissue Culture – Tested for the attribute of cell attachment and growth utilizing an attachment-dependent mammalian cell line. Cell attachment within 24 hours is required for acceptance.

Quality Control Testing – Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass

Packaging Inspection – Pass

Cell Attachment & Growth Treatment Verification - Pass

Lot Number Designation:

8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture; Last 3 digits – Batch identification.