



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Product Description

Catalog Number: 4512-8L

Product Description: Proculture® Flask, Spinner, Glass, 4 Vertical Sidearms, 8000 mL
G.P.I. 400-120 Center Cap, GL-45 Sidearm Caps

Component Materials:

- Glass - Type I Class A - Low-Expansion Borosilicate Glass,
per ASTM E438 (Standard Specification for Glasses in Laboratory Apparatus)
- Center Cap - Polybutylene terephthalate (PBT)
- Liner - Silicone
- SA Cap - Polypropylene (PP)
- Pour Ring - Polypropylene (PP)

Attributes:

- Capacity - 8000 mL
- Product Dimensions - 234 mm OD x 405 mm High without Cap

Autoclaving:

This product will withstand repeated standard autoclave cycles of 121°C at 15 psi for 15 minutes.

Heating / Cooling:

This product is primarily designed for a specific use that does not involve direct heating. In normal use, temperatures will be from -50° to +125°C. This is below the temperature at which thermal failure could be a problem.

Vacuum / Pressure:

This product is not designed for and is not rated for vacuum or pressure applications.

Sterilization:

Shipped Not Sterile

Lot Number Designation:

Each production lot is uniquely identified with an 8 digit lot number as follows: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Revision Date:

5 AUG 2013

Rev No: 1