



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Corning Incorporated Life Sciences

Registered
ISO 9001:2008

Product Description

Catalog Number: 4680

Product Description: Corning® 96 well high content imaging plate, black with clear 127µm bottom, tissue culture treated, with lid

Component Materials:

- | | | |
|--------------|---|--|
| Plate walls | - | Cyclic Olefin Co-polymer, meets <i>USP, Class VI</i> requirements for plastic containers and closures. Black cyclic olefin co-polymer concentrate. |
| Plate bottom | - | Cyclic Olefin Co-polymer, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Lid | - | Virgin Polystyrene, meets <i>USP, Class VI</i> requirements for plastic containers and closures. |
| Gasket | - | Polyester film |

Product Dimensions:

- | | | | | | |
|--------------------------|---|--------------|---------------------------|---|----------|
| Length of Plate | - | 5.030 in. | Diameter of Well @ top | - | .197 in. |
| Width of Plate | - | 3.365 in. | Diameter of Well @ bottom | - | .177 in. |
| Depth of Well | - | .452 in. | Height with Lid | - | .650 in. |
| Tolerances of Dimensions | - | +/- .010 in. | Total well volume | - | 200 µL |

Sterilization:

This lot has been irradiated and dosimetrically released based on ANSI/AAMI/ISO 11137 *Sterilization of healthcare products-Requirements for validation and routine control-Radiation sterilization*.

Sterility Assurance Level: SAL 10⁻³

Surface Characterization:

Surface is characterized to be hydrophilic and negatively charged, composed of 9-17% oxygen atoms. This surface composition has been optimized for cell attachment and growth.

Cell Attachment and Growth Characteristics:

The product has been tested for the attribute of cell attachment and growth using an attachment-dependent mammalian cell line in serum supplemented media.

Optical Characteristics:

The product is made of opaque black cyclic olefin walls to minimize well to well crosstalk and background fluorescence and /or luminescence. The bottom is made of clear cyclic olefin co-polymer to permit direct microscopic viewing and imaging.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

- | | |
|-------------------------|---|
| Visual Attributes: | Visual and microscopic examination of the product. |
| Packaging: | Inspection for seal and barrier integrity, accurate labeling and correct product configuration. |
| Flatness: | Measurement for interwell flatness. |
| Cell Culture Treatment: | Wettability test using water to insure the presence of a hydrophilic surface. |

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian Date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 1