



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Catalog Number: 9459

Product Description: Costar® 1 x 8 Stripwell™, high binding EIA/RIA plate, flat bottom, without lid, red

Component Materials:

- Strips - Virgin Polystyrene, meets *USP Class VI requirements for plastic containers and closures.*
- Holder - Polystyrene, Heavy metal free (meets *CONEG* req.) color concentrate.
- Foil - Red Hot Stamp Foil

Product Dimensions:

- | | | | | | |
|--------------------------|---|--------------|---------------------------|---|----------|
| Length of Plate | - | 5.025 in. | Diameter of Well @ Top | - | .270 in. |
| Width of Plate | - | 3.370 in. | Diameter of Well @ Bottom | - | .250 in. |
| Depth of Well | - | .420 in. | Height | - | .560 in. |
| Tolerances of Dimensions | - | +/- .010 in. | Working Volume per well | - | .36mL |

BSE/TSE:

Product complies with the latest revision of EMA/410/01 "Note for Guidance on minimizing the risk of transmitting animal spongiform encephalopathy agents via human and veterinary medicinal products" by virtue of all bovine derived material having been processed per specific conditions of section 6.4 of EMA/410/01.

Surface Characterization:

Surface is characterized to be high binding for EIA/RIA tests. Binding capacity is approximately 500 ng IgG/cm² . cm².

Optical Characterization:

The product is made of clear polystyrene to allow background absorbance to be \pm .005 optical density units from the mean.

Binding Characteristics:

Each lot has been tested for the attribute of well to well protein binding utilizing a modified ELISA to measure precision and percent low and high wells from the mean. Coefficient of variation must be \leq 3% and high and low wells \leq 8% from the mean. Product has a Costar Certified Surface Chemistry.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

- | | |
|--------------------|--|
| Visual Attributes: | Visual and microscopic examination of the product. |
| Packaging: | Inspection for accurate labeling, and correct product configuration. |
| Functional: | Tested for protein binding precision. |

Lot Number Designation:

8 Digit Lot Number: First 3 digits – Julian date, start of manufacturing; Next 2 digits – Year of manufacture; Last 3 digits – Batch identification.