



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 2592

Product Description: Costar® 1 x 8 Stripwell™, high binding EIA/RIA plate, flat bottom, without lid

Component Materials:

Strips - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

Holder - Polystyrene, Heavy metal free (meets *CONEG* req.) color concentrate.

Product Dimensions:

Length of Plate	- 5.025 in.	Diameter of Well @ top	- .270 in.
Width of Plate	- 3.370 in.	Diameter of Well @ bottom	- .250 in.
Depth of Well	- .420 in.	Height	- .560 in.
Tolerances of Dimensions	- +/- .010 in.	Working Volume per well	- .36 mL

Surface Characterization:

Surface is characterized to be high binding for EIA/RIA tests. Binding capacity is approximately 500 ng IgG/cm². This lot has been gamma irradiated at a specific dose range that has been validated to render the surface high binding for proteins.

Optical Characteristics:

The product is made of clear polystyrene to allow background absorbance to be \pm .005 optical density units from the mean.

Binding Characteristics:

Each lot has been tested for the attribute of well to well protein binding utilizing a modified ELISA to measure precision and percent low and high wells from the mean. Coefficient of variation must be \leq 3% and high and low wells \leq 8% from the mean. Product has a Costar Certified Surface Chemistry.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes:

Visual and microscopic examination of the product.

Packaging:

Inspection for accurate labeling, and correct product configuration.

Functional:

Tested for protein binding precision.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5