



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3207

Product Description: Costar® 1.7 mL Prelubricated Microcentrifuge Tube

Component Materials:

Tube - Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

| | | | | | |
|------------|---|-----------|--------------------|---|-------|
| Length | - | 1.535" | I.D. @ top of tube | - | .362" |
| Tolerances | - | +/- .050" | O.D. @ cap | - | .500" |

Special Feature:

The tube is autoclavable.

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Pyrogens:

The product has been tested and has met the criteria established in the Guidelines on Validation of the Limulus Amebocyte Lysate Test as an End-Product Endotoxin Test for Human and *Animal* Parenteral Drugs, Biological Products and Medical Devices (FDA, December, 1987).

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Visual Attributes: Visual examination of the product.

Packaging: Accurate labeling, and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 5