



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3635

Product Description: Corning **@96-well UV-Plate, without lid**

Component Materials:

- Plate walls - Virgin Acrylic, meets *USP, Class VI* requirements for plastic containers and closures.
- Plate bottom - Flexible Fluorinated-chlorinated Thermoplastic, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

- | | | | | | |
|--------------------------|---|--------------|---------------------------|---|-------------|
| Length of Plate | - | 5.030 in. | Diameter of Well @ top | - | .270 in. |
| Width of Plate | - | 3.365 in. | Diameter of Well @ bottom | - | .250 in. |
| Depth of Well | - | .420 in. | Height | - | .560 in. |
| Tolerances of Dimensions | - | +/- .010 in. | Volume | - | 370 μ L |

Optical Characteristics:

The bottom is made of UV-transparent material, and is molded to the acrylic top for greater strength and maximum leak resistance. The mean background O.D. @ 260 nm and 280 nm is ≤ 0.1 .

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

- | | |
|--------------------|--|
| Visual Attributes: | Visual examination of the product. |
| Packaging: | Inspection for accurate labeling, and correct product configuration. |
| Integrity: | Fluid leakage and crosstalk for 24 hours using a colored solution. |
| UV Absorbance: | Optical density at 260 nm and 280 nm. |

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian Date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 9