



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**Corning Incorporated
Life Sciences**

**Registered
ISO 9001:2008**

Product Description

Catalog Number: 3898

Product Description: Costar ® 96-well assay plate, v-bottom, without lid

Component Materials:

Plate - Virgin Polystyrene, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

| | | | | | |
|-----------------------------|---|--------------|---------------------------|---|--------------|
| Length of Plate | - | 5.035 in. | Diameter of Well @ top | - | .270 in. |
| Width of Plate | - | 3.370 in. | Diameter of Well @ bottom | - | .250 in. |
| Depth to the bottom of well | - | .420 in. | Height | - | .560 in. |
| Tolerances of Dimensions | - | +/- .010 in. | Working Volume per well | - | 0.1 - 0.2 mL |

Surface Characterization:

Surface is characterized to be hydrophilic and negatively charged for solution based assay.

Optical Characteristics:

The product is made of clear polystyrene to allow transparency.

Bovine Spongiform Encephalopathy and Transmissible Spongiform Encephalopathy:

This product is manufactured with animal free materials.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Surface Treatment: Wettability test using water to insure the presence of a hydrophilic surface

Visual Attributes: Visual and microscopic examination of the product.

Packaging: Inspection for accurate labeling, and correct product configuration.

Rev No: 7