



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Corning Incorporated
Life Sciences

Registered
ISO 9001:2008

Product Description

Catalog Number: 8163

Product Description: Costar® Spin-X® centrifuge tube filter cellulose acetate membrane 0.45 µm pore size non-sterile

Component Materials:

- Tube - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Filter Unit - Virgin Polypropylene, meets *USP, Class VI* requirements for plastic containers and closures.
- Membrane - Cellulose Acetate, 0.45 µm pore size, meets *USP, Class VI* requirements for plastic containers and closures.

Product Dimensions:

- | | | | | | |
|-------------------------|---|----------|---------------------------|---|--------------|
| Filter Unit Length | - | .808 in. | Tube Length w/filter unit | - | 1.800 in. |
| Filter Unit ID/top | - | .380 in. | Tube OD below shoulder | - | .434 in. |
| Filter Unit OD/shoulder | - | .375 in. | Tolerances | - | +/- .050 in. |
| Working Volume | - | .5 mL | Tube Volume | - | 2 mL |

Maximum RCF:

16,000

Membrane Bubble Point Integrity:

The membrane has been tested according to an established procedure to determine the water bubble point. The membrane meets the established criteria of minimum 26 psi.

RNase/DNase Testing:

This product has been tested and is free of any detectable RNase/DNase contamination.

Performance Testing:

Each manufacturing lot is sampled and tested in accordance with Standard Operating Procedures.

Integrity Testing: Forward pressurization of filter unit to 10 psi.

Visual Attributes: Visual examination of the product.

Packaging: Inspection for accurate labeling and correct product configuration.

Lot Number Designation:

8 Digit Lot Number: First 3 digits - Julian date, start of manufacturing; Next 2 digits - Year of manufacture; Last 3 digits - Batch identification.

Rev No: 7