



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

**(FOR RESEARCH USE ONLY. DO NOT USE IT IN CLINICAL DIAGNOSIS !)**

## **Human Parainfluenza Virus IgM ELISA Kit**

Catalog No: E-HD-E108

96T/96T\*2

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: [techsupport@elabscience.com](mailto:techsupport@elabscience.com)

Website: [www.vetassay-elab.com](http://www.vetassay-elab.com)

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

**Test principle**

This ELISA kit uses Indirect-ELISA as the method to detect the HPIV-IgM in human serum (plasma). The ELISA Microtiter plate provided in this kit has been pre-coated with Mouse-anti-human IgM ( $\mu$  chain). Samples are added to the ELISA Microtiter plate wells and the HPIV-IgM in which will be captured. Free components are washed away. The HRP conjugated HPIV antigen is added to each well and react with the compound to form “anti- $\mu$  chain – HPIV IgM –HRP conjugated HPIV antigen” compound. The TMB substrate is added to initiate the color developing reaction. The presence of HPIV-IgM can be determined according to the OD value after colorimetric assay with the Micro-plate Reader.

**Kit components**

Item	Specification
ELISA Microtiter plate	96 wells
Positive Control	1 mL
Negative Control	1 mL
HRP Conjugated Working Solution	12 mL
Sample Diluent	12 mL
20×Concentrated Wash Buffer	50 mL
Substrate Reagent A	6 mL
Substrate Reagent B	6 mL
Stop Solution	6 mL
Plate Sealer	3 pieces
Sealed Bag	1
Manual	1 copy

**Experimental instrument**

Micro-plate Reader with 450 nm wavelength filter or dual-wavelength (450/630 nm)

High-precision transferpettor, EP tubes and disposable pipette tips

37°C Incubator or water bath

Deionized water

Absorbent paper

Loading slot for Wash Buffer

## Requirements of sample

1. Human serum can be used as detected sample. Fresh collected serum samples should be fully centrifuged, then take clear liquid for test. The suspended fibrous protein may cause a false positive result if not fully precipitated. Avoid of samples with hyperlipidemia (triglyceride  $\geq 20$  g/L), hemolysis (hemoglobin  $\geq 10$  g/L) or jaundice (bilirubin  $\geq 0.2$  g/L). Obviously contaminated samples can't be detected.
2. Do not use heated inactivated samples. Heat inactivation will degrade antibodies.
3. Samples can be stored at 2~8°C for one week. If samples not tested in a week, store them at -20°C and avoid freeze-thaw cycles.
4. Bring all reagents to room temperature (18~25°C) for more than 30 min before use. Freezing samples should be mixed fully before test.

## Assay procedure

Bring all reagents to room temperature for 30 min. Dilute the 20×Concentrated Wash Buffer for 20 times with distilled water.

### 1. Add sample:

- (1) Take out Micro-plate and mark it, reserve 1 well for blank control (empty), 3 wells for negative control, 2 wells for positive control (100  $\mu$ L of control serum for each well). (Blank well is not necessary for dual-wavelength detection)
  - (2) Dilute the tested **Serum** with **Sample Diluent** at 1:10 into sample well (add 100  $\mu$ L of sample diluent and add 10  $\mu$ L of serum sample), mix fully.
  - (3) Gently tap the plate to ensure thorough mixing.
2. **Incubate:** Cover the ELISA plate with sealer. Incubate for 30 min at 37°C.
  3. **Wash:** After incubation, remove the plate sealer and aspirate the liquid of each well. Repeat the washing procedure for 5 times with wash buffer and immerse for 30-60 sec each time.
  4. **HRP conjugate:** Add 100  $\mu$ L of HRP Conjugate Working Solution to each well except the blank control well.
  5. **Incubate:** Cover the ELISA plate with sealer. Incubate for 30 min at 37°C.
  6. **Wash:** After incubation, remove the plate sealer and aspirate the liquid of each well. Repeat the washing procedure for 5 times with wash buffer and immerse for 30-60 sec each time.
  7. **Add substrate:** Add 50  $\mu$ L of Substrate Reagent A and 50  $\mu$ L of Substrate Reagent B to each well. Gently tap the plate to ensure thorough mixing. Cover with a new plate sealer. Incubate for 15 min at 37°C in dark.
  8. **Stop reaction:** Add 50  $\mu$ L of Stop Solution to each well, gently tap the plate to ensure thorough mixing.
  9. **OD Measurement:** Set the micro-plate reader wavelength at 450 nm (it is recommended to set the dual wavelength at 450 nm/630 nm) to detect A value of each well. Blank well is not needed when using dual wavelength 450 nm/630 nm for detection.

## Reference value

### 1. Result analysis

- (1) Use each test result independently. Determine the result according to the Cut Off value.
- (2) Calculate the Cut Off:  $\text{Cut Off(C.O)} = 0.10 + \text{negative control(NC) average A value}$  (when NC average  $A_{450} < 0.05$ , calculate at 0.05; while NC average  $A_{450} \geq 0.05$ , calculate at the actual value).

### 2. Quality control

- (1) Blank well (just chromogenic agent and stop solution) absorbance  $\leq 0.08$ .
- (2) Positive control (PC)  $A_{450} > 0.80$ .
- (3) Negative control (NC)  $A_{450} < 0.10$ .

The experimental result is valid if quality control is valid.

### 3. Determination of results

- (1) Positive result: Sample absorbance  $\geq$  Cut Off.
- (2) Negative result: Sample absorbance  $<$  Cut Off.

## Interpretation of results

1. Negative result indicates there is no HPIV-IgM antibody detected in samples, while positive result means the opposite.
2. The positive result of HPIV-IgM antibody is an important index of HPIV infection.
3. The high prevalence of positive serum was associated with repeated infection of different subtypes of HPIV in the early stage of childhood. Make comparison with two samples collected at intervals of 8-10 days, if the results showed that there was a significant increase in antibody concentration (such as an increase of 4 times) in the latter blood sample, which indicated an acute infection of HPIV.

## Limitations of test method

The detection results of this kit are only for clinical reference. For confirmation of the diagnosis, please combine the symptoms and other methods of detection, this detection cannot be used as the only criteria for clinical diagnosis.

## Notes

1. Wear gloves and work clothes during experiment, and the disinfection and isolation system should be strictly executed. All the waste should be handled as contaminant.
2. The Stop Solution is corrosive, it should be avoided to contact with skin and clothing. Wash immediately with plenty of water if contact it carelessly.
3. The ELISA plate obtained from cold storage conditions should be adjusted to room temperature before use. The unused plate should be kept in a sealed bag with desiccant.
4. 20×Concentrated Wash Buffer at low temperature condition is easy to crystallize, it should be adjusted to room temperature in order to dissolve completely before use.
5. Each well must be filled with liquid when washing in order to prevent residual free enzyme.
6. The tested sample should be kept fresh.
7. The results shall depend on the readings of the Micro-plate Reader.
8. Do not use components from different batches of kit.

## Storage and shelf life

Store unopened at 2 to 8°C for 12 months. Do not freeze.

Please store the opened kit at 2~8°C, protect from light and moisture. The shelf life of the opened kit is up to 1 months.

**Expiry date:** expiration date is on the box.