



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic)

DON (Deoxynivalenol) Lateral Flow Assay Kit

Catalog No: E-TO-C003

20T/50T/80T

Version Number:	V1.2
Replace version:	V1.1
Revision Date:	2024.03.14

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.

Test principle

This kit uses the principle of Immunochromatography assay for the qualitative detection. It can detect Deoxynivalenol (DON) in samples, such as grain, feed, etc. After adding the sample solution into the sample well of detection card, DON in the sample solution combine with the gold-labelled antibody, so as to prevent the combining of gold-labelled antibody with DON conjugate on the cellulose membrane. When the concentration of DON in the sample solution is more than the detection limit, the detect line do not show color reaction and the result is positive. When the concentration of DON in the sample solution is less than the detection limit, the detect line shows color and the result is negative.

Technical indicator

Detection limit: Grain, Feed ---500 ppb; Oil ---250 ppb

Kits components

Item	Specifications
Detection card (with disposable dropper)	20/50/80T/kit
Manual	1 copy

Other materials required but not supplied

Instruments: Homogenizer, Oscillators, Centrifuge, Graduated pipette, Balance (sensibility 0.01 g)

High-precision transferpettor: Single channel (20-200 μ L, 100-1000 μ L)

Reagent: N-hexane

Notes

1. FOR RESEARCH USE ONLY. Do not use product out of date or in a broken aluminum foil.
2. The detection card should be adjusted to room temperature after removed from the refrigerator before opening. The opening detection card should be used as soon as possible so as not to be invalid because of moisture.
3. Avoid of contacting the white membrane at the middle of the sample well.
4. The disposable dropper cannot be mixing to avoid the cross-contaminant.
5. The tested sample should be clear, no turbidity particle and no bacterial pollution, otherwise it is easy to result in abnormal phenomena such as obstruction, unobvious color, etc., which affect the judgment of the experiment result.
6. If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.
7. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.
8. Each reagent is optimized for use in the E-TO-C003. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-TO-C003 with different lot numbers.

Storage and expiry date

Storage: Store at 2-30°C. With cool and dry environment.

Expiry date: expiration date is on the packing box.

Sample pretreatment

Restore all reagents and samples to room temperature before use.

1. Sample pretreatment Notice:

Experimental apparatus should be clean, and the disposable dropper should be disposable to avoid the experiment result be interfered by the contamination.

2. Sample pretreatment procedure:

2.1 Pretreatment of grain, feed:

- (1) Take 1±0.05 g of crushed homogenate to a 50 mL centrifuge tube, add 10 mL of deionized water. Oscillate for 5 min and centrifuge at 4000 r/min for 10 min at room temperature.
- (2) Take 60 µL of the supernatant to analysis.

Note: Detection limit: 500 ppb

Dilute methods of different detection limit:

Crushed sample	1 g	1 g	1 g	1 g
Deionized water	5 mL	10 mL	20 mL	40 mL
Dilution factor	5	10	20	40
Detection limit	250 ppb	500 ppb	1 ppm	2 ppm

(**Note:** add 20 mL of water into 1g sample first, then oscillate and centrifuge, take 1mL of the supernatant and adding 1ml of deionized water if the sample need to be diluted for 40 multiple.)

2.2 Pretreatment of oil (vegetable oil, sesame oil, salad oil, peanut oil):

- (1) Take 1±0.05 g of sample to a 50 mL centrifuge tube, add 4 mL of **N-hexane** and 5 mL of deionized water. Oscillate for 5 min and centrifuge at 4000 r/min for 10 min at room temperature.
- (2) Discard the supernatant, take 60 µL of the lower layer liquid to analysis.

Note: Detection limit: 250 ppb

Experiment procedure

1. Tear the aluminum foil bag of the detection card and take out the detection card, and put it on a smooth, clean table.
2. Take the prepared clear sample supernatant with the matching disposable dropper, add 2-3 drops (about 60 µL) of sample to the sample well (S) vertically and slowly (Avoid foaming).
3. Incubate for 8 to 10 minutes and then judge the results immediately.

Judgment of result

1. **Negative:** The test line region (T) and the control line region (C) show a line at the same time in the observation well. It indicates the content of DON in the sample is lower than detection limit or the sample doesn't contain DON.
2. **Positive:** Only the control line region (C) show a line in the observation well. It indicates the content of DON in the sample is higher than detection limit.
3. **Invalid:** The control line region (C) does not show a line in the observation well. It indicates operation process is wrong or the test card is invalid.

