

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere Liefer- und Versandbedingungen

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

linkedin.com/company/szaboscandic in





OTA (Ochratoxin A) Lateral Flow Assay Kit

Catalog No: E-TO-C011

20T/50T/80T

Version Number: V1.2
Replace version: V1.1

Revision Date: 2024.03.14

This manual must be read attentively and completely before using this product.

If you have any problems, please contact our Technical Service Center for help.

Toll-free: 1-888-852-8623 Tel: 1-832-243-6086 Fax: 1-832-243-6017

Email: techsupport@elabscience.com

Website: www.elabscience.com

Please kindly provide us the lot number (on the outside of the box) of the kit for more efficient service.



Test principle

This kit uses the principle of Immunochromatography assay for the qualitative detection. It can detect OTA (Ochratoxin A) in samples, such as grain, feed, etc. After adding the sample solution into the sample well of detection card, OTA in the sample solution combine with the gold-labelled antibody, so as to prevent the combining of gold-labelled antibody with OTA conjugate on the cellulose membrane. When the concentration of OTA in the sample solution is more than the detection limit, the detect line do not show color (or shows lighter color than control line) and the result is positive. When the concentration of OTA in the sample solution is less than the detection limit, the detect line show color (shows equal or darker color than control line) and the result is negative.

Technical indicator

Detection limit: Grain, Feed---5ppb.

Kits components

Item	Specifications
Detection Card	20/50/80 T/kit
Reconstitution Buffer	1 vial
Manual	1 copy

Note: All reagent bottle caps must be tightened to prevent evaporation and microbial pollution

Other materials required but not supplied

Instruments: Homogenizer, Nitrogen Evaporators, Water bath, Oscillators, Centrifuge, Graduated

pipette, Balance (sensibility 0.01g).

High-precision transferpettor: Single channel (20-200 μL, 100-1000 μL)

Reagent: Methanol



Notes

- 1. FOR RESEARCH USE ONLY. Do not use product out of date or in a broken aluminum foil.
- 2. The detection card should be adjusted to room temperature after removed from the refrigerator before opening. The opening detection card should be used as soon as possible so as not to be invalid because of moisture.
- 3. Avoid of contacting the white membrane at the middle of the sample well.
- 4. The disposable dropper cannot be mixing to avoid the cross-contaminant.
- 5. The tested sample should be clear, no turbidity particle and no bacterial pollution, otherwise it is easy to result in abnormal phenomena such as obstruction, unobvious color, etc., which affect the judgment of the experiment result.
- 6. If the samples are not indicated in the manual, a preliminary experiment to determine the validity of the kit is necessary.
- 7. The kit is used for rapid screening of actual samples. If the test result is positive, the instrument method such as HPLC, LC/MS, etc. can be used for quantitative confirmation.
- 8. Each reagent is optimized for use in the E-TO-C011. Do not substitute reagents from any other manufacturer into the test kit. Do not combine reagents from other E-TO-C011 with different lot numbers.

Storage and valid period

Storage: Store at 2-30°C. With cool and dry environment.

Expiry date: expiration date is on the packing box.

Sample pretreatment

Restore all reagents and samples to room temperature before use

1. Sample pretreatment Notice:

Experimental apparatus should be clean, and the disposable dropper should be disposable to avoid the experiment result be interfered by the contamination.

2. Sample pretreatment procedure

2.1 Pretreatment of grain (rice, wheat, corn, etc.), feed:

- (1) Weigh 2 ± 0.05 g of crushed sample into 15 mL centrifuge tube, add 3 mL of **Methanol**, and oscillate for 3 min.
- (2) Centrifuge at 4000 r/min at room temperature for 3 min.
- (3) Take 1 mL of supernatant to another centrifuge tube (liquid has been layered), dry in nitrogen evaporators/water bath at 50-60 °C. (Part of the yellow grease on the bottom layer is normal. Please do it in a ventilated environment.)
- (4) Dissolve the residual with 0.3 mL of **Reconstitution Buffer**, oscillate for 30 s.

Note: Detection limit: 5 ppb



Experiment procedure

- 1. Tear the aluminum foil bag of the detection card and take out the detection card, and put it on a smooth, clean table.
- 2. Take the prepared sample with the matching disposable dropper, add 120 μ L of sample solution to the gold-labelled micro well, wait for 2 min, and whip the purple residual with a burette until it is completely dissolved (Avoid foaming). Wait for 2 min again, remove all the liquid of the gold-labelled micro well into the sample well, count down at the same time.
- 3. Incubate for 5 to 8 minutes and then judge the results immediately.

Judgment of result

- 1. **Negative:** The control line region (C) show color, the test line region (T) shows equal or darker than line C. It indicates the content of OTA in the sample is lower than detection limit or the sample doesn't contain OTA.
- 2. **Positive:** The control line region (C) show color, the test line region (T) shows no color or lighter color than line C. It indicates the content of OTA in the sample is higher than detection limit.
- 3. **Invalid:** The control line region (C) shows no color. It indicates operation process is wrong or the test card is invalid.

