



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Troponin I Control

HumaFIA SR

Package Size 2 x 0.6 mL

REF 16090/36

IVD

Intended Purpose

Troponin I Control HumaFIA SR is a control to check the accuracy and reliability of the Troponin I HumaFIA SR assay.

The control is designed to be used together with the Troponin I HumaFIA SR reagent on semi-automated HumaFIA analyzer.

For professional in vitro diagnostic use only.

Reagents and Contents

CTR1	0.6 mL	Control Level 1 (green), sealed in foil pouch Lyophilized control, spiked with Troponin I
CTR2	0.6 mL	Control Level 2 (brown), sealed in foil pouch Lyophilized control, spiked with Troponin I
DIL	1.4 mL	Diluent Buffer

The **CTR** concentrations levels are lot specific, see target value sheet.

Applicable system

REF	
16090	HumaFIA
16090/35	Troponin I HumaFIA SR

Storage/Stability

Unopened at 2...8°C	Up to the stated expiration date
Opened pouch at 18-28°C	Reconstitution should be completed within 30 mins
Reconstituted at 2...8°C	5 days open vial

Store upright and protected from light.

Procedure

Follow the procedure exactly as described.

Procedural Notes

- Do not use **CTR** after the expiration date.
- Bring **CTR** to room temperature (18-28°C) before use.
- Gently invert the reconstituted **CTR** at least 10 times avoiding the formation of foam.
- Treat the control in the same manner as patient specimen in the assay procedure. Test should be completed at 22-24°C.
- Read the control results, refer to the target value sheet.
- Do not return remaining **CTR** to the original vial once used.

Reconstitution

- Before reconstitution, gently tap the vial to ensure the lyophilized material is at the bottom of the vial.
- Before reconstitution, first remove carefully the cap to prevent the contents from ejection.
- Reconstitute **CTR** by adding exactly 0.6 mL **DIL**.
- Gently invert the reconstituted **CTR** at least 10 times, avoiding the formation of foam. Check all **CTR** has dissolved.
- Close the cap and keep **CTR** at room temperature for 15 minutes. After 15 minutes, coat all surfaces of the bottle by rotating and inverting the bottle. Continue mixing until the solution is homogeneous and all lyophilized material is reconstituted.
- Use the necessary volume of **CTR**.
- Close the cap and store the remaining volume at 2-8°C.

Notes

- This product is not intended to be used as standard material.

Assay Procedure

Refer to the instruction for the use of Troponin I HumaFIA SR

REF 16090/35.

Safety Notes

All patient specimens, calibrators and controls should be handled as potentially infectious. All materials of animal origin avoid many risks associated with the use of human serum (e.g. Hepatitis B and C, HIV). Nevertheless, all material of human or animal origin should still be treated as potentially infectious material.

For users in the European Union only: Please report any serious incident that has occurred in relation to the device to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.

P234 Keep only in original container.

P260 Do not breathe dust/fume/gas/mist/vapors/spray.

P262 Do not get in eyes, on skin, or on clothing.

P281 Use personal protective equipment as required.

P303+P361+P353 IF ON SKIN (or hair): Take off immediately all contaminated clothing. Rinse skin with water/shower.

P305+P351+P338 IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing.

P337+P313 If eye irritation persists: Get medical advice/attention.

P401 Store in accordance with local/regional/national/international regulations.

P501 Dispose of contents/container in accordance with local/regional/ national/international regulations.



Beijing HumaDX Tech Co., Ltd.

Address: Room 1309, 3rd Floor, 1st Block, No. 55 Jiachuang 2nd Road, BDA 101111, Beijing, China

E-mail: service@humadx.com.cn Tel: +86-10-80828658



HUMAN Gesellschaft für Biochemica und Diagnostica mbH

Address: Max-Planck-Ring 21, 65205, Wiesbaden, Germany
Tel: +49-6122-99880, E-Mail: human@human.de



Effective Date: Feb 7, 2024
Version: HD/CE 36 202402 V001