



# SZABO SCANDIC

Part of Europa Biosite

## Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!  
See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

[mail@szabo-scandic.com](mailto:mail@szabo-scandic.com)

[www.szabo-scandic.com](http://www.szabo-scandic.com)

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 



# FLU A&B & COVID-19 & RSV & Adeno

## Antigen Combo Test Cassette

### Instructions For Use

**R224T020B0C0 20 Tests/kit**    **R224T025B0C0 25 Tests/kit**

FOR PROFESSIONAL IN VITRO DIAGNOSTIC USE ONLY

#### INTENDED USE

The FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette is a rapid chromatographic immunoassay for the qualitative detection of influenza A (FLU A), influenza B (FLU B), COVID-19 (SARS-CoV-2), respiratory syncytial virus (RSV) and adenovirus (Adeno) antigen in nasopharyngeal swab specimens. This test is intended for use as an aid in the differential diagnosis of influenza A, influenza B, COVID-19, respiratory syncytial virus and adenovirus viral infections in humans in conjunction with clinical and epidemiological risk factors.

#### Summary

The FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette is a rapid chromatographic immunoassay for the qualitative detection of influenza A, influenza B, COVID-19, respiratory syncytial virus (RSV) and adenovirus antigen in nasopharyngeal swab specimens from individuals with suspected influenza A&B/COVID-19/RSV/adenovirus infection in conjunction with clinical presentation and the results of other laboratory tests. Results are for the detection of influenza A+B, COVID-19, RSV and adenovirus antigen. An antigen is generally detectable in upper respiratory specimens during the acute phase of infection. Positive results indicate the presence of viral antigen, but clinical correlation with patient medical history and other diagnostic information is necessary to determine infection status. Positive results do not rule out other bacterial/viral infection. Negative results should be treated as a presumption and confirmed with a molecular assay, if necessary for patient management. Negative results should be considered in the context of a patient's recent exposures, medical history and the presence of clinical signs and symptoms consistent with influenza A+B, COVID-19, RSV and adenovirus. The FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette is intended for use by trained clinical laboratory personnel.

#### TEST PRINCIPLE

The FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette is a qualitative membrane strip based immunoassay for the detection of influenza A virus, and influenza B virus, COVID-19 virus, respiratory syncytial virus (RSV) and adenovirus antigen in nasopharyngeal swab specimens.

In this test procedure, influenza A antibody, influenza B antibody, COVID-19-N antibody, RSV virus antibody and adenovirus antibody is immobilized in the different test line regions of the device. After a nasopharyngeal swab specimen is placed in the specimen well, it reacts with influenza A antibody, influenza B antibody, COVID-19-N antibody, respiratory syncytial virus antibody and adenovirus antibody coated particles that have been applied to the specimen pad. This mixture migrates chromatographically along the length of the test strip and interacts with the immobilized influenza A antibody, influenza B antibody, COVID-19-N antibody, respiratory syncytial virus and adenovirus antibody. If the specimen contains influenza A virus antigen, influenza B virus antigen, COVID-19 virus antigen, respiratory syncytial virus antigen, adenovirus antigen, a colored line will appear in the corresponding test line region, indicating a positive result. If the specimen does not contain influenza A virus antigen, influenza B virus antigen, COVID-19 virus antigen, respiratory syncytial virus antigen, adenovirus antigen, a colored line will not appear in these regions, indicating a negative result. To serve as a procedural control, a colored line will always appear at the control line region, indicating that proper volume of specimen has been added and membrane wicking has occurred.

#### MATERIALS

##### Material Provided

Kit size: 20 tests/kit

Test Cassette - 20

Instructions for use - 1

Workstation - 1

Disposable swab - 20

Extraction tube with buffer solution - 20

Kit size: 25 tests/kit

Test Cassette - 25

Instructions for use - 1

Workstation - 1

Disposable swab - 25

Extraction tube with buffer solution - 25

##### Materials not provided:

Timer, specimen collection container, centrifuge, disposable latex gloves, sealed bag and disinfectant.

#### STORAGE AND STABILITY

- Stored at 2-30°C, the validity period of the product is 24 months. Do not freeze.
- The test cassette should be used within 1 hour after opening the sealed pouch. If the temperature is higher than 30°C or in high humidity environment, it should be to use immediately.
- Kit contents are stable until the expiration date printed on the package.
- Keep away from sunlight, moisture and heat.

#### WARNINGS AND PRECAUTIONS

1. For *in vitro* diagnostic use only. Do not use after expiration date. Do not reuse the test cassette.
2. Do not eat, drink, or smoke in the area where the samples or kits are handled.
3. Handle all samples as if they contain infectious agents.
4. Observe established precautions against microbiological hazards throughout all procedures and follow the standard procedures for proper disposal of samples.
5. The test cassette should remain in a sealed foil pouch until use. Do not use the test cassette if the pouch is damaged or opened.
6. Allow the test device and specimens to equilibrate at room temperature (15-30°C) and humidity (<80%) prior to testing.
7. After use, the test cassette can be disposed of with household waste in a sealed bag.
8. Follow local standard biosafety guidelines for handling and disposal of potential infective materials.

#### SPECIMEN COLLECTION AND HANDLING

- Only the swab provided in the kit is to be used for nasopharyngeal swab collection.
- To collect a nasopharyngeal swab sample, carefully insert the swab into the nostril exhibiting the most visible drainage, or the nostril that is most congested if drainage is not visible. Using gentle rotation, push the swab until resistance is met at the level of the turbinates (less than one inch into the nostril). Rotate the swab 5 times or more against the nasopharyngeal wall and then slowly remove from the nostril. Using the same swab, repeat sample collection in the other nostril.
- Testing should be performed immediately after specimen collection. Do not leave the specimens at room temperature for prolonged periods.
- Bring specimens to room temperature prior to testing.
- If it is not possible to test immediately, it is strongly recommended that the swab is placed in a clean, unused plastic extraction tube labeled with patient information to maintain best performance and avoid possible contamination. The sample can be kept tightly sealed in this extraction tube at room temperature (15-30°C or 59-86°F) for a maximum of one hour. Make sure that the swab is firmly seated in the extraction tube and that the cap is tightly closed. If a delay of more than one hour occurs, discard the sample. A new sample must be taken for the test.
- If specimens are to be transported, they should be packaged according to local regulations for the transport of pathogens.

#### TEST PROCEDURE


Please read the instructions for use carefully before use.

Allow the test cassette, buffer to room temperature 15-30 °C (59-86 °F) before testing.

1. Place the extraction tube in the workstation. Peel off aluminum foil seal from the top of the extraction tube containing the extraction buffer.
2. Place the swab into the extraction tube. Rotate the swab for 10-15 seconds while pressing the head against the inside of the tube to release the specimens in the swab.
3. Remove the swab while squeezing the swab head against the inside of the extraction tube as you remove it to expel as much liquid as possible from the swab. Discard the swab in accordance with your biohazard waste disposal protocol.
4. When ready to test, open the pouch at the notch and take out the test cassette. Place the test cassette on a clean, flat surface.
5. Fit the tube tip or close the cap onto the tube, then invert the extraction tube and add 3 drops of specimen (approximately 90µL) into each of the specimen well (S) and then start the timer.
6. Read the result at 15 minutes. If left unread for 20 minutes or more, the results are invalid and a repeat test is recommended.

##### Notes:

Applying sufficient amount of specimen is essential for a valid test result. If migration (the wetting of membrane) is not observed in the test window after one minute, add one more drop of specimen.



## INTERPRETATION OF RESULTS

A procedural control is included in the test. A colored line appearing in the control line region (C) is considered an internal procedural control. It confirms sufficient sample volume, adequate membrane wicking, and correct procedural technique.


### 1. Interpretation of Flu A/B Results:

**Positive:** Control line and at least one test line appear on the membrane. The appearance of A test line indicates the presence of Flu A antigen. The appearance of B test line indicates the presence of Flu B antigen. And if both A and B line appear, it indicates that the presence of both Flu A and Flu B antigen.

**Negative:** One colored line appears in the control region(C). No apparent colored line appear in the test line region.

**Invalid:** Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test device . If the problem persists, discontinue using the test kit immediately and contact your local distributor.

### Interpretation of Flu A/B Results




### 2. Interpretation of COVID-19/RSV/Adenovirus Results

**Positive:** Two lines appear. One line should always appear in the control line region(C), and another one apparent colored line should appear in the test line region.

**Negative:** One colored line appears in the control region(C). No apparent colored line appear in the test line region.

**Invalid:** Control line fails to appear. Insufficient specimen volume or incorrect procedural techniques are the most likely reasons for control line failure. Review the procedure and repeat the test with a new test device . If the problem persists, discontinue using the test kit immediately and contact your local distributor.

### Interpretation of COVID-19/RSV/Adenovirus Results



**\*NOTE:** The intensity of the color in the test line regions may vary depends on the concentration of virus antigen. Therefore, any shade of color in the test line region should be considered positive.

## PROCEDURAL CONTROL

There are internal procedural controls in the test. A colored line displayed in the control area (C) is an internal procedural control. It confirms the presence of a sufficient amount of sample and correct procedure.

## LIMITATIONS OF THE TEST METHOD

1. This test detects both viable (live) and non-viable, FLU A/B, SARS-CoV, COVID-19, respiratory syncytial virus(RSV) and adenovirus antigen. Test performance depends on the amount of virus (antigen) in the sample and may or may not correlate with viral culture results performed on the same sample .
2. A negative test result may occur if the level of antigen in a sample is below the detection limit of the test.
3. The performance of FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette was evaluated using the procedures provided in this product insert only. Modifications to these procedures may alter the performance of the test.
4. False negative results may occur if a specimen is improperly collected , transported, or handled.

5. False results may occur if specimens are tested past 1 hour of collection. Specimen should be test as quickly as possible after specimen collection .

6. Positive test results do not rule out co-infections with other pathogens

7. Positive test results do not differentiate between SARS-CoV and COVID-19 antigen .

8. Negative test results are not intended to rule in other viral or bacterial infections .

9. Negative results, from patients with symptom onset beyond seven days, should be treated as presumptive and confirmation with a molecular assay, if necessary, for patient management, may be performed.

10. If the differentiation of specific influenza A virus , influenza B virus and COVID-19 virus and respiratory syncytial virus(RSV), and adenovirus antigen is needed, additional testing, in consultation with local public health departments.

11. A negative result does not mean a person is not infectious or does not have influenza. If symptoms persist the person should seek medical attention and further testing if required.

## PERFORMANCE CHARACTERISTICS

### 1. Clinical performance

#### 1.1 The evaluation assay results for Influenza A are as below:

Table 1. Influenza A: Comparison of evaluation assay and reference assay:

Reference		A leading Commercial FLU A test		Total Results
Method	Result	Positive	Negative	
FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette	Positive	98	1	99
	Negative	3	649	652
TOTAL RESULTS		101	650	751

The coincidence rate of sensitivity: 97.03% (95%CI\*: 91.56 %-99.38%)

The coincidence rate of specificity: 99.85% (95%CI\*:99.14%-99.99%)

Total coincidence rate: 99.47% (95%CI\*: 98.64%-99.85%)

#### 1.2 The evaluation assay results for Influenza B are as below:

Table 2. Influenza B: Comparison of evaluation assay and reference assay

Reference		A leading Commercial FLU B test		Total Results
Method	Result	Positive	Negative	
FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette	Positive	101	0	101
	Negative	2	648	650
TOTAL RESULTS		103	648	751

The coincidence rate of sensitivity: 98.06% (95%CI\*: 93.16%-99.76%)

The coincidence rate of specificity: 100.00% (95%CI\*: 99.54%-100.00%)

Total coincidence rate: 99.73% (95%CI\*: 99.04%-99.97%)

#### 1.3 The evaluation assay results for SARS-CoV-2 are as below:

Table 3. SARS-CoV-2: Comparison of evaluation assay and reference assay

Reference		A leading Commercial COVID-19 Test		Total Results
Method	Result	Positive	Negative	
FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette	Positive	132	2	134
	Negative	3	614	617
TOTAL RESULTS		135	616	751

The coincidence rate of sensitivity: 97.78% (95%CI\*: 93.64%-99.54%)

The coincidence rate of specificity: 99.67% (95%CI\*: 98.93%-99.96%)

Total coincidence rate: 99.33% (95%CI\*: 98.45%-99.78%)

Table 4. SARS-CoV-2: Subjects on Days Post-Symptom Onset

Days post Symptom onset	Number of samples	Commercial Test Positive Result	FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette (COVID-19 Test Positive Result)	95% Confidence Interval
1	4	4	4 (100.00%)	47.29%-100.00%
2	10	10	10 (100.00%)	74.11%-100.00%
3	15	15	15 (100.00%)	81.90%-100.00%
4	18	18	18 (100.00%)	84.67%-100.00%
5	23	23	22 (95.65%)	78.05%-99.89%
6	27	27	26 (96.30%)	81.03%-99.91%
7	35	35	34 (97.14%)	85.08%-99.93%
Total	132	132	129 (97.73%)	93.50%-99.52%

1.4 The evaluation assay results for RSV are as below:

Table 5. RSV: Comparison of evaluation assay and reference assay

Reference		A leading Commercial RSV test		Total Results
Method	Result	Positive	Negative	
FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette	Positive	124	6	130
	Negative	3	618	621
TOTAL RESULTS		127	624	751

The coincidence rate of sensitivity: 97.64% (95%CI\*: 93.25%-99.51%)

The coincidence rate of specificity: 99.04% (95%CI\*: 97.92%-99.65%)

Total coincidence rate: 98.80% (95%CI\*: 97.74%-99.45%)

1.5 The evaluation assay results for Adenovirus are as below:

Table 6. Adenovirus: Comparison of evaluation assay and reference assay

Reference		A leading Commercial Adenovirus test		Total Results
Method	Result	Positive	Negative	
FLU A&B & COVID-19 & RSV & Adeno Antigen Combo Test Cassette	Positive	149	2	151
	Negative	3	597	600
TOTAL RESULTS		152	599	751

The coincidence rate of sensitivity: 98.03% (95%CI\*: 94.34%-99.60%)

The coincidence rate of specificity: 99.67% (95%CI\*: 98.80%-99.96%)

Total coincidence rate: 99.33% (95%CI\*:98.45% -99.78%)

2. Interference

For Flu A

The following compounds have been tested using the Flu A test and no interference was observed.

Analytes	Conc.	Analytes	Conc.
Whole Blood	20µL/mL	Oxymetazoline	0.6mg/mL
Mucin	50µg/mL	Phenylephrine	12mg/mL
Budesonide Nasal Spray	200µL/mL	Rebetol	4.5µg/mL
Dexamethasone	0.8mg/mL	Relenza	282ng/mL
Fluni solide	6.8ng/mL	Tamiflu	1.1µg/mL
Mupirocin	12mg/mL	Tobryamycin	2.43mg/mL

For Flu B

The following compounds have been tested using the Flu B test and no interference was observed.

Analytes	Conc.	Analytes	Conc.
Whole Blood	20µL/mL	Oxymetazoline	0.6mg/mL

Mucin	50µg/mL	Phenylephrine	12mg/mL
Budesonide Nasal Spray	200µL/mL	Rebetol	4.5µg/mL
Dexamethasone	0.8mg/mL	Relenza	282ng/mL
Fluni solide	6.8ng/mL	Tamiflu	1.1µg/mL
Mupirocin	12mg/mL	Tobryamycin	2.43mg/mL

For COVID-19

The following compounds have been tested using the COVID-19 test and no interference was observed.

Analytes	Conc.	Analytes	Conc.
Whole Blood	20µL/mL	Oxymetazoline	0.6mg/mL
Mucin	50µg/mL	Phenylephrine	12mg/mL
Budesonide Nasal Spray	20µL/mL	Rebetol	4.5µg/mL
Dexamethasone	0.8mg/mL	Relenza	282ng/mL

For RSV

The following compounds have been tested using the respiratory syncytial virus test and no interference was observed.

Mucin	Meropenem	Lopinavir	Dexamethasone
Blood	Tobramycin	Ritonavir	Flunisolide
α-Interferon	Histamine hydrochloride	Abidol	Triamcinolone Acetonide
Zanamivir	Benfulin	Levofloxacin	Budesonide
Ribavirin	Oxymetazoline	Azithromycin	Momitsone
Oseltamivir	NaCl(including preservative)	Human Anti-mouse Antibody(HAMA)	Fluticasone
Palamivir	Beclomethasone	Ceftriaxone	Biotin

For Adenovirus

The following compounds have been tested using the adenovirus antigen test and no interference was observed.

Mucin	Meropenem	Lopinavir	Dexamethasone
Blood	Tobramycin	Ritonavir	Flunisolide
α-Interferon	Histamine hydrochloride	Abidol	Triamcinolone Acetonide
Zanamivir	Benfulin	Levofloxacin	Budesonide
Ribavirin	Oxymetazoline	Azithromycin	Momitsone
Oseltamivir	NaCl(including preservative)	Human Anti-mouse Antibody(HAMA)	Fluticasone
Palamivir	Beclomethasone	Ceftriaxone	Biotin

3. Cross-reactivity

For Flu A

The Flu A Antigen Test has been tested for other virus(Table below). The results showed no cross-reactivity.

Influenza B	Human Rhinovirus 14	Arcanobacterium	Staphylococcus aureus subsp.aureus
Human coronavirus OC43	Human Rhinovirus 16	Candida albicans	Staphylococcus epidermidis
Coronavirus NL63	Measles	Corynebacterium	Streptococcus pneumoniae
Pseudomonas aeruginosa	Mumps	Escherichia coli	Streptococcus pyogenes
Nesseria subllava	Parainfluenza virus 2	Moraxella catarrhalis	Streptococcus salivarius
Respiratory syncytial virus	Parainfluenza virus 3	Neisseria lactamica	Streptococcus sp group F
Human Rhinovirus 2	COVID-19 virus		

For Flu B

The Flu B Antigen Test has been tested for other virus(Table below). The results showed no cross-reactivity .

Influenza A H3N2	Human Rhinovirus 14	Arcanobacterium	Staphylococcus aureus subsp.aureus
Human coronavirus OC43	Human Rhinovirus 16	Candida albicans	Staphylococcus epidermidis
Coronavirus NL63	Measles	Corynebacterium	Streptococcus pneumoniae
Pseudomonas aeruginosa	Mumps	Escherichia coli	Streptococcus pyogenes
Nesseria subllava	Parainfluenza virus 2	Moraxella catarrhalis	Streptococcus salivarius
Respiratory syncytial virus	Parainfluenza virus 3	Neisseria lactamica	Streptococcus sp group F

Human Rhinovirus 2	COVID-19 virus	Influenza A H1N1	
--------------------	----------------	------------------	--

**For COVID-19**

The COVID-19 Antigen Test has been tested for other virus(Table below). The results showed no cross - reactivity.

Pseudomonas aeruginosa	Human Rhinovirus 14	Arcanobacterium	Staphylococcus aureus subsp.aureus
Human coronavirus OC43	Human Rhinovirus 16	Candida albicans	Staphylococcus epidermidis
Coronavirus NL63	Measles	Corynebacterium	Streptococcus pneumoniae
Influenza A H1N1	Mumps	Escherichia coli	Streptococcus pyogenes
Influenza A H3N2	Parainfluenza virus 2	Moraxella catarrhalis	Streptococcus salivarius
Influenza B	Parainfluenza virus 3	Neisseria lactamica	Streptococcus sp group F
Human Rhinovirus 2	Respiratory syncytial virus	Nesseria subllava	

**For RSV**

The respiratory syncytial virus Antigen Test has been tested for other virus(Table below). The results showed no cross-reactivity.

Nesseria subllava	Human Rhinovirus 14	Arcanobacterium	Staphylococcus aureus subsp.aureus
Human coronavirus OC43	Human Rhinovirus 16	Candida albicans	Staphylococcus epidermidis
Coronavirus NL63	Measles	Corynebacterium	Streptococcus pneumoniae
Influenza A H1N1	Mumps	Escherichia coli	Streptococcus pyogenes
Influenza A H3N2	Parainfluenza virus 2	Moraxella catarrhalis	Streptococcus salivarius
Influenza B	Parainfluenza virus 3	Neisseria lactamica	Streptococcus sp group F
Human Rhinovirus 2	Pseudomonas aeruginosa		

**For Adenovirus**

The adenovirus Antigen Test has been tested for other virus(Table below). The results showed no cross-reactivity.

Pseudomonas aeruginosa	Human Rhinovirus 14	Arcanobacterium	Staphylococcus aureus subsp.aureus
Human coronavirus OC43	Human Rhinovirus 16	Candida albicans	Staphylococcus epidermidis
Coronavirus NL63	Measles	Corynebacterium	Streptococcus pneumoniae
Influenza A H1N1	Mumps	Escherichia coli	Streptococcus pyogenes
Influenza A H3N2	Parainfluenza virus 2	Moraxella catarrhalis	Streptococcus salivarius
Influenza B	Parainfluenza virus 3	Neisseria lactamica	Streptococcus sp group F
Human Rhinovirus 2	Nesseria subllava		

**BIBLIOGRAPJY**

1. Goff J L, Kerneis S, Elie C, et al. Evaluation of saliva molecular point of care for detection of COVID-19 in ambulatory care[J] . Scientific Reports.
2. Slinger R, Milk R, Gaboury I, et al. Evaluation of the Quick Lab RSV Test, a New Rapid Lateral-Flow Immunoassay for Detection of Respiratory Syncytial Virus Antigen[J]. Journal of Clinical Microbiology, 2004, 42(8):3731-3733.
3. Alhethel A, Albarrag A M, Somily A M, et al. H1N1 testing: A comparative study between the rapid BD Directigen EZ Flu A+B test and PCR[J]. Journal of Pure and Applied Microbiology, 2014,8(3):2265-2270.
4. Floriana B, Chiara M, Chiara F, et al. Evaluation of the diagnostic performances of two commercially available assays for the detection of enteric adenovirus antigens[J]. Diagnostic Microbiology & Infectious Disease, 2021, 101(3).
5. Eunkyung S, Amy L, Preeti J, et al. 1138Human Adenovirus (HAdV) Blood Viral Detection is Associated with Higher Viral Load in

**INDEX OF SYMBOLS**

	Consult Instructions for Use
--	------------------------------

	Contains sufficient for <n> tests
--	-----------------------------------

	Temperature limit
--	-------------------

	In vitro diagnostic medical device
	Catalogue number
	CE conformity marking
	Do not use if package is damaged and consult instructions for use

	Use by date
	Batch code
	Authorized representative in the European Community/ European Union

	Manufacturer
	Date of manufacture
	Do not re-use



**Jiangsu Mole Bioscience Co., Ltd.**



**MedUnion S.L.**

G116 Building, No.805, Jiankang Avenue, Medical New&Hi-tech District, 225300 Taizhou, Jiangsu Province, P.R. China  
 info@molechina.com  
 www.molechina.com

Carrer de Tapioles,33, 2-1, Barcelona, 08004, Spain  
 Tel: +34-644173535  
 Email: rep@themedunion.com

Number:301100078200

Reversion Date:2024.10.23

EN version: 0.0