



SZABO SCANDIC

Part of Europa Biosite

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten!
See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere [Liefer- und Versandbedingungen](#)

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

[linkedin.com/company/szaboscandic](https://www.linkedin.com/company/szaboscandic) 

Monosan[®] Plus HRP Polymer anti-Rabbit

REF / Cat. No.: MON-APP202 60 tests, 6 ml

Instructions for use

Intended use

Monosan[®] Plus HRP Polymer anti-Rabbit is designed for the qualitative detection of antigens in fixed paraffin-embedded tissue sections, in frozen tissue sections, and in cytological samples. It is developed for use in combination with mono- and polyclonal primary antibodies and sera obtained from rabbits.

The reagent can be used for examining tissues fixed in different solutions, e.g. formalin (neutrally buffered), B5, Bouin, ethanol, or HOPE.

This product is for research use only, not for drug, diagnostic or other use.

Summary and Explanation

The purpose of the immunohistochemical staining is to make tissue and cell antigens visible.

Monosan[®] Plus HRP Polymer anti-Rabbit is a highly sensitive detection reagent intended for use in immunohistochemistry and immunocytochemistry. The enzyme polymer in this kit consists of several molecules of secondary antibodies covalently bound to several molecules of horse radish peroxidase (HRP).

Visualisation occurs via an enzyme-substrate reaction in the presence of a colourising reagent which permits microscopical analysis.

The reagent is suitable for the detection of mono- and polyclonal primary antibodies and sera obtained from rabbits. In contrast to other detection techniques, which often use the streptavidin-biotin system the Monosan[®] Plus HRP Polymer anti-Rabbit avoids the problem of background staining caused by endogenous biotin in the tissue.

Principle of the method

Paraffin-embedded tissue sections are first deparaffinised and rehydrated.

Endogenous peroxidase activity in the tissue may cause non-specific staining. This enzyme activity can be blocked by incubation with 3% H₂O₂-solution ("Peroxide block").

Background staining caused by unspecific binding of the primary antibody or the secondary antibody in the HRP polymer is minimized by incubation with a protein blocking solution ("Blocking Solution"). This step can be omitted if the primary antibodies are diluted in an appropriate buffer.

The next step is incubation with the specific primary antibody. After washing, the HRP-polymer is applied and incubated. Any excess of unbound HRP-polymer is thoroughly washed away after incubation. The addition of the chromogenic substrate starts the enzymatic reaction of the alkaline phosphatase which leads to colour precipitation where the primary antibody is bound. The colour can be observed with a light microscope.

The chromogen used determines the colour. The chromogen AEC leads to the formation of a red-brown product of reaction at the place of the target antigen. The chromogen DAB forms a dark brown precipitate.

Reagent provided

REF / Cat. No. MON-APP202

6 ml Monosan® Plus HRP-Polymer anti-Rabbit

Substrate systems recommended:**Substrate systems recommended:**

Permanent AEC Kit	Cat. No. MON-APP173	2000 tests
AEC Single Solution	Cat. No. MON-APP169	80 tests
	Cat. No. MON-APP170	1250 tests
AEC Substrate Kit	Cat. No. MON-APP171	500 tests
	Cat. No. MON-APP172	5000 tests
DAB Substrate Kit	Cat. No. MON-APP176	500 tests
	Cat. No. MON-APP177	5000 tests
DAB High Contrast Kit	Cat. No. MON-APP174	500 tests
	Cat. No. MON-APP175	5000 tests

Materials required but not supplied

Positive und negative control tissue

Xylene or suitable substitutes

Ethanol, distilled H₂O

Reagents for enzyme digestion or heat pre-treatment

Wash buffer PBS or TBS (Cat. No. MON-APP165/166)

PAP Pen (Cat. No. MON-APP190)

Primary antibody (user-defined)

Primary antibody diluent (Cat. No. MON-APP148/149)

Negative control reagent

Chromogenic substrate

Counter stain solution

Mounting medium

Cover slips

Storage and handling

The solutions should be stored at 2-8°C without further dilution. Please store the reagents in a dark place and do not freeze them. Under these conditions the solutions are stable up to the expiry date indicated on the label. They should not be used after the expiry date. A positive and a negative control have to be carried out in parallel to the test material. If you observe unusual staining or other deviations from the expected results which could possibly be caused by the kit reagents, please contact Monosans' technical support or your local distributor.

Precautions

Use by qualified personnel only.

Wear protective clothing to avoid eye, skin or mucous membrane contact with the reagents. In case of a reagent coming into contact with a sensitive area, wash the area with large amounts of water.

ProClin 300, used for stabilisation, is not considered hazardous material in the concentration used. Material safety data sheets (MSDS) for the pure substances are available upon request.

Microbial contamination of the reagents must be avoided, since otherwise non-specific staining might appear.

Reagent preparation

- Reagents should be at room temperature when used.
- Deparaffinise and rehydrate paraffin-embedded tissue sections.
- Pre-treatment (optional) with HIER (*Heat Induced Epitope Retrieval*) or enzymatic digestion.
- Tissue sections have to be completely covered with the different reagents in order to avoid drying out.

Staining procedure

- | | |
|---|-------------------|
| 1. Peroxide blocking (3% H ₂ O ₂) | 10 min. |
| 2. Washing with wash buffer | 1 x 2 min. |
| 3. Blocking Solution (This step is optional) | 5 min. |
| 4. Washing with wash buffer | 1 x 2 min. |
| 5. Primary antibody (optimally diluted) or negative control reagent | 30-60 min. |
| 6. Washing with wash buffer | 3 x 2 min. |
| 7. HRP-polymer anti Rabbit | 30 min. |
| 8. Washing with wash buffer | 3 x 2 min. |
| 9. AEC or DAB
(Controlling the colour intensity via light microscope is recommended.) | 5-15 min. |
| 10. Stopping the reaction with distilled H ₂ O when the desired colour intensity is attained | |
| 11. Counterstaining and blueing | |
| 12. Mounting: aqueous with AEC, permanent with DAB or Permanent AEC | |

Quality control

We recommend carrying out a positive and a negative control with every staining run. The positive control permits the validation of appropriate processing of the sample. If the negative control has a positive result, this points to unspecific staining.

Expected results

During the reaction of the substrate with alkaline phosphatase in the presence of a chromogen, a coloured precipitate is formed at the location of the bound primary antibody. This reaction only takes place if the target antigen is existent in the tissue. The chromogen used determines the colour of the precipitate. The analysis is carried out using a light microscope.

Limitations of the procedure

Immunohistochemistry is a complex method in which histological as well as immunological detection methods are combined. Tissue processing and handling prior to immunostaining, for example variations in fixation and embedding or the inherent nature of the tissue can cause inconsistent results (Nadji and Morales, 1983).

Inadequate counterstaining and mounting can influence the interpretation of the results. The colour intensity of the reaction product can decrease with time, especially when exposed to light.

Monosan® guarantees that the product will meet all requirements described from its shipping date until its expiry date, as long as the product is correctly stored and utilized. No additional guarantees can be given. Under no circumstances shall Monosan® be liable for any damages arising out of the use of the reagent provided.

Explanation of the symbols on the product label:

	Bestellnummer Catalog Number Reference du catalogue		Chargenbezeichnung Batch Code Code du lot		Reizend Irritant Irritant
	Gesundheitsschädlich Harmful Nocif		Giftig Toxic Toxique	 Hersteller / Manufacturer / Fabricant Monosan® Frontstraat 2c 5405 PB Uden The Netherlands Tel: (+31) 413 251115 Fax: (+31) 413 266605 info@monosan.com www.monosan.com	
	Verwendbar bis Use By Utiliser jusque				
	Gebrauchsanweisung beachten Consult Instructions for use Consulter les instructions d'utilisation		Lagerungstemperatur Temperature Limitation Limites de température		

Troubleshooting

If you observe unusual staining or other deviations from the expected results which could possibly be caused by the reagents, please read these instructions carefully, contact Monosans' technical support or your local distributor.

No staining on an actually positive control slide:

1. Reagents were not used in the proper order.
2. Chromogenic substrate solution was too old.
3. Bleaching because chromogen and mounting medium are incompatible.
4. The antigen/epitope in the tissue was insufficiently accessible to the primary antibody. Try a pre-treatment such as heat pre-treatment or enzyme digestion. If you used a pre-treatment it should be extended.
5. Primary antibody not from rabbit, but from a different species.
6. The antigen/epitope was not stable in the fixation and/or pre-treatment procedure used. Try another fixation or pre-treatment.

Weak staining:

1. Inadequate fixation or overfixation.
2. Incomplete deparaffinisation.
3. The antigen/epitope in the tissue was insufficiently accessible to the primary antibody. If you used heat pre-treatment or enzyme digestion it should be extended.
4. Excessive incubation with Blocking Solution or insufficient washing after this step.
5. Too much wash buffer remains on the slides after washing, diluting the reagents applied in the next step.
6. If you are using PBS-based wash buffer: the activity of alkaline phosphatase in the reagents is blocked if too much wash buffer remains on the slides.
7. Incubation times were too short or primary antibody concentration too low.
8. Chromogenic substrate solution was too old.

Non-specific background staining or overstaining:

1. Incomplete deparaffinisation.
2. Excessive tissue adhesive on slides.
3. Insufficient washing especially after the incubation with the enzyme polymer or the chromogenic substrate solution. These washings are critical.
4. Tissue was allowed to (partially) dry out with reagents on.
5. Unspecific binding of the primary antibody. Please use a Blocking Solution or dilute the primary antibody in appropriate diluents.
6. Incubation time of the primary antibody was too long or primary antibody concentration too high.
7. Incubation time of the chromogenic substrate solution was too long or reaction temperature too high (e.g. if temperature in the laboratory is high).

Performance characteristics

Monosan[®] has conducted studies to evaluate the performance of the kit reagents. The product has been found to be suitable for the intended use.

Bibliography

Elias JM "Immunohistopathology – A practical Approach to Diagnosis" ASCP Press 2003

Nadji M and Morales AR. Ann N.Y. Acad Sci 420:134-9, 1983

FOR RESEARCH USE ONLY, NOT FOR DRUG, DIAGNOSTIC OR OTHER USE.

Explanation of the symbols on the product label:

	Bestellnummer Catalog Number Reference du catalogue		Chargenbezeichnung Batch Code Code du lot		Reizend Irritant Irritant
	Gesundheitsschädlich Harmful Nocif		Giftig Toxic Toxique	 Hersteller / Manufacturer / Fabricant Monosan® Frontstraat 2c 5405 PB Uden The Netherlands Tel: (+31) 413 251115 Fax: (+31) 413 266605 info@monosan.com www.monosan.com	
	Verwendbar bis Use By Utiliser jusque				
	Gebrauchsanweisung beachten Consult Instructions for use Consulter les instructions d'utilisation		Lagerungstemperatur Temperature Limitation Limites de température		