

# Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere Liefer- und Versandbedingungen

## Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

linkedin.com/company/szaboscandic in





### Instructions For Use

## A00152-C-IFU-RUO

Rev. Date: June 5, 2017

**Revision: 1** 

Page 1 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

# Cytokeratin, Pan; Clones AE-1 & AE-3 (Concentrate)

Availability/Contents: <u>Item #</u> <u>Volume</u>

A00152-C.1 0.1 ml A00152-C 1 ml

**Description:** 

Species: Mouse

Immunogen: Human epidermal keratin

Clone: AE-1 & AE-3

Isotype: IgG1, kappa (AE-1); IgG1, kappa (AE-3)

Entrez Gene ID: 3848 (CK1); 3850 (CK3); 3851 (CK4); 3852 (CK5); 3853 (CK6A); 3856 (CK8); 3858 (CK10);

3861 (CK14); 3866 (CK15); 3868 (CK16); 3880 (CK19)

Hu Chromosome Loc.: 12q13.13 (CK1); 12q13.13 (CK3); 12q13.13 (CK4); 12q13.13 (CK5); 12q13.13 (CK6); 12q13.13

(CK8); 17q21.2 (CK10); 17q21.2 (CK14); 17q21.2 (CK15); 17q21.2 (CK16); 17q21.2 (CK19)

Synonyms: K1B; KRT1B; Keratin, type II cytoskeletal 1b; K77; CK-1B; Keratin 1B; Keratin-77; Cytokeratin-

1B; Type-II Keratin Kb39

Mol. Weight of Antigen: 40-67kDa

Format: 200µg/ml of Ab purified from Bioreactor Concentrate by Protein A/G. Prepared in 10mM PBS

with 0.05% BSA & 0.05% azide.

Specificity: This antibody cocktail recognizes acidic (Type I or LMW) and basic (Type II or HMW)

cytokeratins, which include 67kDa (CK1); 64kDa (CK3); 59kDa (CK4); 58kDa (CK5); 56kDa (CK6); 52kDa (CK8); 56.5kDa (CK10); 50kDa (CK14); 50kDa (CK15); 48kDa (CK16); 40kDa (CK19). This antibody stains cytokeratins present in normal and abnormal human tissues and

has shown high sensitivity in the recognition of epithelial cells and carcinomas.

Background: Twenty human keratins are resolved with two-dimensional gel electrophoresis into acidic (pl

<5.7) and basic (pl >6.0) subfamilies. Many studies have shown the usefulness of keratins as markers in cancer research and tumor diagnosis. AE-1 & AE-3 is a broad spectrum anti pancytokeratin antibody cocktail which differentiates epithelial tumors from non-epithelial tumors, e.g. squamous vs. adenocarcinoma of the lung, liver carcinoma, breast cancer, and esophageal

cancer. It has been used to characterize the source of various neoplasms and study the distribution of cytokeratin-containing cells in epithelia during normal development and during the

development of epithelial neoplasms.

Species Reactivity: Human, Monkey, Cow, Dog, Rabbit, Mouse, Rat, Chicken. Others not known.

Positive Control: Skin, Adeno- or Squamous carcinomas.

Cellular Localization: Cytoplasmic

Titer/ Working Dilution: Immunohistochemistry (Frozen and Formalin-fixed): 0.5-1 µg/ml

Flow Cytometry: 0.5-1 µg/million cells

Immunofluorescence:  $1-2 \mu g/ml$  Western Blotting:  $0.5-1 \mu g/ml$ 

Microbiological State: This product is not sterile.

15 00 C

ScyTek Laboratories, Inc. 205 South 600 West

Logan, UT 84321 U.S.A. CE

Emergo Europe
Prinsessegracht 20

2514 AP The Hague, The Netherlands

Doc: IFU-Template2-8rev2



# Instructions For Use

# A00152-C-IFU-RUO

Rev. Date: June 5, 2017

**Revision: 1** 

Page 2 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

Uses/Limitations: Not to be taken internally.

For Research Use Only.

This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded

tissue sections, to be viewed by light


microscopy.

Do not use if reagent becomes cloudy.

Do not use past expiration date.

Non-Sterile.

Ordering Information and Current Pricing at www.scytek.com



FFPE skin stained with Cytokeratin, Pan; Clones AE-1 & AE-3.

### Procedure:

- 1. Tissue Section Pretreatment (Required): Staining of formalin fixed, paraffin embedded tissue sections is enhanced by pretreatment with Citrate Plus (ScyTek catalog# CPL500).
- 2. **Primary Antibody Incubation Time:** We suggest an incubation period of 30 minutes at room temperature. However, depending upon the fixation conditions and the staining system employed, optimal incubation should be determined by the user.
- 3. Visualization: For maximum staining intensity we recommend the "UltraTek HRP Anti-Polyvalent Lab Pack" (ScyTek catalog# UHP125, see IFU for instructions) combined with the "DAB Chromogen/Substrate Bulk Pack (High Contrast)" (ScyTek catalog# ACV500, see IFU for instructions).

**Precautions:** Contains Sodium Azide as a preservative (0.09% w/v).

Do not pipette by mouth.

Avoid contact of reagents and specimens with skin and mucous membranes.

Avoid microbial contamination of reagents or increased nonspecific staining may occur.

This product contains no hazardous material at a reportable concentration according to U.S. 29 CFR 1910.1200,

OSHA Hazardous Communication Standard and EC Directive 91/155/EC.

#### References:

- Woodock-Mitchell J et. al. Journal of Cell Biology 1982;95:580-8.
- Tseng SCG et. al. Cell 1982; 30361.

Warranty: No products or "Instructions For Use (IFU)" are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our IFU or website. Our warranty is limited to the actual price paid for the product. ScyTek Laboratories, Inc. is not liable for any property damage, personal injury, time or effort or economic loss caused by our products. Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue specimen may cause variations in results. Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used.

Storage: 2° C

ScyTek Laboratories, Inc. 205 South 600 West Logan, UT 84321 U.S.A.

FC REP Emergo Europe Prinsessegracht 20 2514 AP The Hague, The Netherlands