

### Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



### Lieferung & Zahlungsart

siehe unsere Liefer- und Versandbedingungen

### Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

### SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

linkedin.com/company/szaboscandic in





### Instructions For Use

## RA0030-C.5-IFU-RUO

Rev. Date: Sept. 23, 2014

**Revision: 1** 

Page 1 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

# Caldesmon, HMW (h-Caldesmon) (Smooth Muscle Marker); Clone CALD1/820 & h-CALD

(Concentrate)

Availability/Contents: <u>Item #</u> <u>Volume</u>
RA0030-C.5 0.5 ml

**Description:** 

Species: Mouse

Immunogen: Recombinant human CALD1 protein (CALD1/820); Crude human uterus extract (h-CALD)

Clone: CALD1/820 & h-CALD

Isotype: IgG1, kappa (CALD1/820); IgG1, kappa (h-CALD)

Entrez Gene ID: 800 (Human)

Hu Chromosome Loc.: 7q33

Synonyms: CAD; CALD1; Caldesmon 1 Isoform 1; Caldesmon 1 Isoform 2; Caldesmon 1 Isoform 3;

Caldesmon 1 Isoform 4; Caldesmon 1 Isoform 5; CDM; HCAD; LCAD; NAG22

Mol. Weight of Antigen: 150kDa

Format: 200µg/ml of Ab purified from Bioreactor Concentrate by Protein A/G. Prepared in 10mM PBS

with 0.05% BSA & 0.05% azide.

Specificity: Recognizes a protein of 150kDa, which is identified as the high molecular weight variant of

Caldesmon. This MAb recognizes only the 150kDa variant (h-Caldesmon) in Western blots of human aortic media extracts and is unreactive with fibroblast extracts from cultivated human

foreskin.

Background: Two closely related variants of human Caldesmon have been identified which are different in

their electrophoretic mobility and cellular distribution. The h-Caldesmon variant (120–150kDa) is predominantly expressed in smooth muscle whereas I-Caldesmon (70–80kDa) is found in non-muscle tissue and cells. Neither of the two variants has been detected in skeletal muscle. Caldesmon is a developmentally regulated protein involved in smooth muscle and non-muscle

contraction.

Species Reactivity: Human. Others not known.

Positive Control: Uterus
Cellular Localization: Cytoplasmic

Titer/ Working Dilution: Immunohistochemistry (Frozen and Formalin-fixed): 0.5-1 µg/ml

Flow Cytometry: 0.5-1 µg/million cells

Immunofluorescence: 1-2 μg/ml Western Blotting: 0.5-1 μg/ml

Immunoprecipitation: 1-2 μg/500μg protein lysate

Microbiological State: This product is not sterile.

Storage: 2° C - 8° C

ScyTek Laboratories, Inc. 205 South 600 West Logan, UT 84321

CE

EC | REP | EmergoEurope (31)(0) 70 345-8570 | Molsnstraat 15

U.S.A. 2513 BH Hague, The Netherlands



# Instructions For Use RA0030-C.5-IFU-RUO

Rev. Date: Sept. 23, 2014

**Revision: 1** 

Page 2 of 2

P.O. Box 3286 - Logan, Utah 84323, U.S.A. - Tel. (800) 729-8350 - Tel. (435) 755-9848 - Fax (435) 755-0015 - www.scytek.com

Uses/Limitations: Not to be taken internally.

For Research Use Only.

This product is intended for qualitative immunohistochemistry with normal and neoplastic formalin-fixed, paraffin-embedded

tissue sections, to be viewed by light

microscopy.

Do not use if reagent becomes cloudy.

Do not use past expiration date.

Non-Sterile.

Ordering Information and Current Pricing at www.scytek.com

### Procedure:

- 1. **Tissue Section Pretreatment (Required):** Staining of formalin fixed, paraffin embedded tissue sections is significantly enhanced by pretreatment with EDTA Buffer (10X) HIER Solution (pH 8.0) (ScyTek catalog# ETA).
- Primary Antibody Incubation Time: We suggest an incubation period of 30 minutes at room temperature.
   However, depending upon the fixation conditions and the staining system employed, optimal incubation should be determined by the user.
- 3. **Visualization:** For maximum staining intensity we recommend the "UltraTek HRP Anti-Polyvalent Lab Pack" (ScyTek catalog# UHP125, see IFU for instructions) combined with the "DAB Chromogen/Substrate Bulk Pack (High Contrast)" (ScyTek catalog# ACV500, see IFU for instructions).

### **Precautions:**

Contains Sodium Azide as a preservative (0.09% w/v).

Do not pipette by mouth.

Avoid contact of reagents and specimens with skin and mucous membranes.

Avoid microbial contamination of reagents or increased nonspecific staining may occur.

This product contains no hazardous material at a reportable concentration according to U.S. 29 CFR 1910.1200,

OSHA Hazardous Communication Standard and EC Directive 91/155/EC.

#### References:

- 1. Watanabe, K., Tajino, T., Sekiguchi, M. and Suzuki, T. 2000. H-Caldesmon as a specific marker for smooth muscle tumors. Comparison with other smooth muscle markers in bone tumors. Am. J. Clin. Pathol. 113: 663-668.
- Frid MG, et al. Phenotypic changes of human smooth muscle cells during development: Late expression of heavy caldesmon and calponin. Dev Biol 1992; 153:185.

### Warranty:

No products or "Instructions For Use (IFU)" are to be construed as a recommendation for use in violation of any patents. We make no representations, warranties or assurances as to the accuracy or completeness of information provided on our IFU or website. Our warranty is limited to the actual price paid for the product. ScyTek Laboratories, Inc. is not liable for any property damage, personal injury, time or effort or economic loss caused by our products. Immunohistochemistry is a complex technique involving both histological and immunological detection methods. Tissue processing and handling prior to immunostaining can cause inconsistent results. Variations in fixation and embedding or the inherent nature of the tissue specimen may cause variations in results. Endogenous peroxidase activity or pseudoperoxidase activity in erythrocytes and endogenous biotin may cause non-specific staining depending on detection system used.

Storage: 2° C

ScyTek Laboratories, Inc. 205 South 600 West Logan, UT 84321 U.S.A. CE

Ec REP EmergoEurope (31)(0) 70 345-8570 Molsnstraat 15 2513 BH Hague, The Netherlands