

Produktinformation



Forschungsprodukte & Biochemikalien



Zellkultur & Verbrauchsmaterial



Diagnostik & molekulare Diagnostik



Laborgeräte & Service

Weitere Information auf den folgenden Seiten! See the following pages for more information!



Lieferung & Zahlungsart

siehe unsere Liefer- und Versandbedingungen

Zuschläge

- Mindermengenzuschlag
- Trockeneiszuschlag
- Gefahrgutzuschlag
- Expressversand

SZABO-SCANDIC HandelsgmbH

Quellenstraße 110, A-1100 Wien

T. +43(0)1 489 3961-0

F. +43(0)1 489 3961-7

mail@szabo-scandic.com

www.szabo-scandic.com

linkedin.com/company/szaboscandic in





INSTRUCTION MANUAL

Quest 5-hmC™ DNA ELISA Kit

Catalog Nos. **D5425 & D5426**

Highlights

- Sensitive and specific quantitation of 5-hydroxymethylcytosine (5-hmC) DNA from a variety of samples.
- Ideal for global 5-hmC detection, tissue-specific 5-hmC quantitation, high-throughput compound screening, and more.
- Streamlined workflow can be completed in as little as 3 hours.

Contents

| Product Contents | 1 |
|-----------------------------|---|
| Specifications | 1 |
| Product Description | 2 |
| Experimental Considerations | 3 |
| Buffer Preparation | 3 |
| Protocol | 4 |
| Appendix | 5 |
| Ordering Information | 6 |
| Related Products | |

For Research Use Only Ver. 1.7.0

Satisfaction of all Zymo Research products is guaranteed. If you should be dissatisfied with this product please call 1-888-882-9682.

Product Contents

| Quest 5-hmC™ DNA ELISA Kit | D5425 (1 × 96 wells) | D5426 (2 × 96 wells) | Storage Temperature |
|--|-----------------------------|-----------------------------|------------------------|
| Coating Buffer | 15 ml | 30 ml | 4 °C |
| 10X ELISA Buffer | 30 ml | 60 ml | 4 °C |
| Anti-5-Hydroxymethylcytosine Polyclonal Antibody (1 mg/ml) | 25 µl | 50 μl | -20 °C |
| Anti-DNA HRP Antibody (100X) | 100 µl | 200 µl | -80 °C |
| HRP Developer | 15 ml | 30 ml | 4 °C |
| Control DNA Set (5 Controls) | 5 × 40 µl | 5 × 40 µl | -20 °C |
| 96-well ELISA Plate (12 x 8-well Strips) | 1 plate | 2 plates | Room Temp. |
| Protocol | 1 | 1 | - |

Note- Integrity of kit components is guaranteed for up to six (6) months from date of purchase. Reagents are routinely tested on a lot-to-lot basis to ensure they provide the highest performance and reliability.

Specifications


- **Sample Sources** This ELISA procedure has been optimized for the detection of 5-hmC in purified genomic DNA that is intact, sheared or fragmented in PBS, Tris-EDTA, or water. The product is also compatible with DNA from other sources.
- **Detection** This system is highly sensitive for 5-hmC DNA and has a lower detection of 0.02% per 100 ng input DNA, making it easy to detect even the smallest percentage of 5-hmC in DNA samples.
- **Equipment Required** An incubator and plate reader (with 405 nm detection) are required. A multi-channel pipettor is recommended. An automated plate washer may be used for blocking and wash steps due to the one-buffer system.

Note - ™ Trademarks of Zymo Research Corporation. This product is for research use only and should only be used by trained professionals. It is not intended for use in diagnostic procedures. Some reagents included with this kit are irritants. Wear protective gloves and eye protection. Follow the safety guidelines and rules enacted by your research institution or facility.


Product Description

The 5-methylcytosine (5-mC) modification of DNA in epigenetic regulation has been well studied over the last several decades. However, the role of the so-called "sixth base", 5-hydroxymethylcytosine (5-hmC), has yet to be defined. Recent studies have associated 5-hmC patterns to transcriptional regulation of genes, which may contribute to normal and disease states of organisms.

The Quest 5-hmC™ DNA ELISA Kit is both sensitive and specific and can be used to accurately detect 5-hmC DNA in a variety of samples. The kit is compatible with a wide range of input DNA including intact vertebrate, plant, and microbial genomic DNA, as well as enzyme-digested and mechanically sheared fragments. The Control DNA Set included with this kit has been calibrated to accurately quantify the percent 5-hmC in sample DNA by use of a standard curve. Also, the fast, streamlined workflow is ideal when analyzing/screening large numbers of samples.



Quest 5-hmC[™] DNA ELISA Kit is a sandwich-based ELISA format. First, Anti-5-Hydroxymethylcytosine Polyclonal Antibody (anti-5-hmC pAb) is coated to the bottom of a well. Single stranded 5-hmC-containing DNA binds to anti-5-hmC pAb which is then recognized by Anti-DNA HRP Antibody. Addition of HRP Developer will produce a greenish-blue color in the wells containing 5-hmC DNA.



The Quest 5-hmC™ DNA ELISA Kit can be used to detect 5-hmC in numerous DNA samples with high specificity as evidenced by comparison with LC-MS/MS-MRM analysis. 5-hmC pAb (200 ng/well) was used to quantitate the amount 5-hmC in 100 ng of single-stranded DNA. For this, % 5-hmC was calculated from a standard curve generated using the Control DNA Set. The figure shows a correlation between the % 5-hmC in DNA samples calculated using the Quest 5-hmC™ DNA ELISA Kit (ELISA) and mass spectrometry (MS).

For **Technical Assistance**, please contact Zymo Research Technical Support at: 1-888-882-9682 or e-mail tech@zymoresearch.com.

Experimental Considerations

All DNA used with the kit must be denatured prior to use. The protocol is optimized for the detection of 5-hmC in 100 ng of denatured (single-stranded) DNA/well. Depending on your experimental design, the amount of input DNA can range from 25-200 ng/well. The Control DNA should be assayed at the same concentration as the sample DNA.

The Control DNA Set consists of five double stranded genomic DNA controls containing a specified percentage of 5-hmC. Each control is provided at a concentration of 100 ng/µl. For 5-hmC detection, not all controls have to be used. For example: Control A (0%) can serve as a negative control, and Control E (0.55%) as a positive control. However, for accurate quantification of 5-hmC percentage, a standard curve must be generated using all controls (see Appendix A, page 5).

Buffer Preparation and Storage

- ✓ Prepare the 1X ELISA Buffer, pH 7.4, by diluting the 10X ELISA Buffer solution (1:10) in deionized water. The 1X ELISA Buffer may be prepared all at once and stored at 4°C for use within one week, or aliquotted and stored at -20°C for up to six months. Repeated freeze/thaw cycles should be avoided.
- ✓ The Coating Buffer, pH 9.6, is ready for use and is stable at room temperature or 4 °C for extended periods of time.
- ✓ The HRP Developer is also ready for use and should be stored at 4°C. For more rapid color development, bring HRP Developer to room temperature before adding to the wells of the ELISA plate.
- ✓ Anti-DNA HRP Antibody can be stored at -20°C for 1 week. For long term storage, the antibody should be kept at -80°C. Avoid freeze/thaw cycles.

Protocol*

It is recommended that samples and controls be assayed in **duplicate** for accurate 5-hmC detection in DNA.

Coating:

- 1. Remove the amount of 8-well strips required to assay samples and standards¹.
- 2. Dilute Anti-5-Hydroxymethylcytosine Polyclonal Antibody (1 mg/ml) to 1 ng/µl in Coating Buffer.
- 3. Add 100 µl/well of the diluted anti-5-hmC pAb². Cover the plate with foil to prevent evaporation and incubate at 37 °C for 1 hour.

Blocking:

- 1. Discard buffer from the wells of the plate. Wash each well with 200 µl of **1X ELISA Buffer** and remove liquid from each well by tapping out excess liquid onto a paper towel. Repeat this wash step 2 more times.
- 2. Add 200 µl of **1X ELISA Buffer** to each well. Cover the plate with foil and incubate at 37 °C for 30 minutes.

DNA Addition/Binding:

- 1. Add 100 ng of each DNA to a PCR tube and bring the final volume to 100 μl with 1X ELISA Buffer³.
 - a. For example, add 1 µl Control A (100 ng/µl) to 99 µl of 1X ELISA Buffer.
- 2. Denature the DNA at 98°C for 5 minutes in a thermocycler. Immediately transfer samples to ice for 10 minutes.
- 3. Discard buffer from the wells of the plate. Remove all liquid from each well by tapping out excess liquid onto a paper towel.
- 4. Add 100 μl of the denatured sample and control DNAs to each well⁵. Cover the plate with foil and incubate at 37 °C for 1 hour.

Addition Anti-DNA HRP Antibody:

- 1. Discard buffer from the wells of the plate. Wash each well with 200 µl of **1X ELISA Buffer**. Remove all liquid from each well by tapping out excess liquid onto a paper towel. Repeat this wash step 2 more times.
- 2. Dilute Anti-DNA HRP Antibody (100X) in 1X ELISA Buffer to final 1X.

 For Example: Add 20 μl of Anti-DNA HRP Antibody 2 ml 1X ELISA Buffer.

 This is enough antibody mix for 20 wells.
- 3. Add 100 μ l of antibody mix to each well. Cover the plate with foil and incubate at 37°C for 30 minutes.

Color Development:

- 1. Discard buffer from the wells of the plate. Wash each well with 200 µl of **1X ELISA Buffer**. Remove all liquid from each well by tapping out excess liquid onto a paper towel. Repeat this wash step 2 more times.
- 2. Add 100 µl of **HRP Developer** to each well and allow color to develop at room temperature for 10 to 60 minutes.
- 3. Use an ELISA plate reader to measure the well absorbance at 405 nm.

Notes:

- * For more information regarding 5-hmC detection and quantification using the **Control DNA Set** refer to Appendix A, page 5.
- ¹The strips of wells that are not used should be stored in a clean, dry, dark place for use at a later date.
- ² Adding 100 µl anti-5-hmC pAb diluted to 1 ng/µl yields 100 ng per well; however 50-400 ng/well 5-hmC pAb can be used to coat wells depending on the DNA sample being detected.
- ³ Do not exceed 100 μl final volume in each PCR tube. If denaturing duplicate Control and sample DNAs in same tube (recommended), bring the final volume to 200 μl after denaturing step.
- ⁵Adding 100 μl of 1 ng/ μl DNA yields a final amount of 100 ng /well; however, 25-200 ng /well DNA can be used with this assay.

Notes:

¹ **Control DNA** should always be included together with the samples for every assay to ensure proper functionality.

Appendix A – Generation of a Standard Curve with the Control DNA Set

For 5-hmC <u>Detection</u>:

Relative levels of 5-hmC in DNA can be determined by comparing the absorbance of samples to **Control A** (0%) serving as a negative control and **Control E** (0.55%) as a positive control¹. Since the percent 5-hmC content is provided for all controls (Table 1, below), any of the **Controls** can be included to approximate the relative levels of 5-hmC in DNA.


For 5-hmC Quantification:

To quantitate the 5-hmC percentage in a DNA sample, a standard curve¹ must be generated using all the provided **Controls**. Plot the **Control** data as absorbance (Y-axis) vs. percent 5-hmC (X-axis) and use the linear regression (equation below) to determine the "% 5-hmC" for the DNA samples (unknowns).

% 5-hmC = (absorbance – y-intercept) Slope

Table 1. Controls 1-5 and corresponding percent (%) 5-hydroxymethylcytosine.

| % 5-hmC | | |
|---------|--|--|
| 0 % | | |
| 0.03 % | | |
| 0.12 % | | |
| 0.23 % | | |
| 0.55 % | | |
| | | |



An example of a standard curve generated with the Control DNA Set. A standard curve was constructed from the absorbance (405 nm) values of Controls A-E (Table 1). The % 5-hmC in any samples is calculated using the equation of the line as shown above.

Ordering Information

| Product Description | Catalog No. | Kit Size |
|----------------------------|----------------|------------------------------|
| Quest 5-hmC™ DNA ELISA Kit | D5425 D5426 | 1 x 96 wells 2 x 96 wells |

| For Individual Sale | Catalog No. | Amount |
|---|----------------------------|------------------|
| Coating Buffer | D5425-1-15 D5425-1-30 | 15 ml 30 ml |
| 10X ELISA Buffer | D5425-2-30 D5425-2-60 | 30 ml 60 ml |
| Anti-5-Hydroxymethylcytosine Polyclonal Antibody (1 mg/ml) | A4001-25 A4001-50 | 25 μl 50 μl |
| Anti-DNA HRP Antibody (100X) | D5425-3-100 D5425-3-200 | 100 μl 200 μl |
| HRP Developer | D5425-4-15 D5425-4-30 | 15 ml 30 ml |
| Control DNA Set | D5425-5-C | 5 × 40 μl |
| Control A (100 ng/µl) | D5425-5-1 | 40 µl |
| Control B (100 ng/μl) | D5425-5-2 | 40 µl |
| Control C (100 ng/µl) | D5425-5-3 | 40 µl |
| Control D (100 ng/µl) | D5425-5-4 | 40 µl |
| Control E (100 ng/μl) | D5425-5-5 | 40 µl |
| 96-well ELISA Plate (12 x 8-well Strips) | C2020 | 1 plate |

Related Products for 5-hmC Analysis:

| Product Name | Size | Catalog No. |
|---|-------------|-------------|
| Quest 5-hmC™ DNA Enrichment Kit | 25 Preps. | D5420 |
| | 50 Preps. | D5421 |
| Quest 5-hmC Detection Kit™ | 25 Preps. | D5410 |
| | 50 Preps. | D5411 |
| | 25 Preps. | D5415 |
| Quest 5-hmC Detection Kit™-Lite | 50 Preps. | D5416 |
| | 50 Rxns. | E2050 |
| Quest <i>Taq</i> ™ PreMix | 200 Rxns. | E2051 |
| Human Matched DNA Set | 2 x 5 μg | D5018 |
| Mouse 5hmC & 5mC DNA Set | 4 x 5 μg | D5019 |
| 5-Methylcytosine & 5-Hydroxymethylcytosine DNA Standard Set | 3 x 2 µg | D5405 |
| DNA Degrades IM | 500 units | E2016 |
| DNA Degradase™ | 2,000 units | E2017 |
| DNA Degradase Plus™ | 250 units | E2020 |
| | 1,000 units | E2021 |
| 5-hmC Glucosyltransferase | 100 units | E2026 |
| | 200 units | E2027 |
| 5-Hydroxymethyl dCTP [100 mM] | 10 µmol | D1045 |
| 5-Hydroxymethylcytosine dNTP Mix [10 mM] | 2.5 µmol | D1040 |
| 5-Methyl dCTP [10 mM] | 1 µmol | D1035 |
| 5-Methylcytosine dNTP Mix [10 mM] | 2.5 µmol | D1030 |

Additional Products for Epigenetics Research:

| Product Name | Size | Catalog No. |
|--|-------------------------------|----------------------------------|
| OneStep qMethyl™ Kit | 1 x 96 | D5310 |
| OneStep qMethyl™-Lite | 1 x 96 | D5311 |
| Zymo <i>Taq</i> ™ DNA Polymerase | 50 200 | E2001 E2002 |
| Zymo <i>Taq</i> ™ PreMix | 50 200 | E2003 E2004 |
| EZ DNA Methylation™ Kit | 50 200 2 x 96 2 x 96 | D5001 D5002 D5003 D5004 |
| EZ DNA Methylation-Gold™ Kit | 50 200 2 x 96 2 x 96 | D5005 D5006 D5007 D5008 |
| EZ DNA Methylation-Direct™ Kit | 50 200 2 x 96 2 x 96 | D5020 D5021 D5022 D5023 |
| EZ DNA Methylation-Startup™ Kit | 1 Kit | D5024 |
| EZ Bisulfite DNA Clean-up Kit™ | 50 200 2 x 96 2 x 96 | D5025 D5026 D5027 D5028 |
| Universal Methylated DNA Standard | 1 set | D5010 |
| Universal Methylated Human DNA Standard | 1 set | D5011 |
| Universal Methylated Mouse DNA Standard | 1 set | D5012 |
| Human HCT116 DKO Methylation Standards | 1 set | D5014 |
| Human HCT116 DKO Non-methylated DNA Standard | 5 µg | D5014-1 |
| Human HCT116 DKO Methylated DNA Standard | 5 µg | D5014-2 |
| Bisulfite Converted Universal Methylated Human DNA Standard | 1 set | D5015 |
| E. coli Non-methylated Genomic DNA | 5 µg | D5016 |
| Methylated-DNA IP Kit | 10 | D5101 |
| ChIP DNA Clean & Concentrator™ | 50 50 | D5201 D5205 |
| Anti-5-Methylcytosine Monoclonal Antibody (clone 10G4) | 50 μg 200 μg | A3001-50 A3001-200 |
| CpG Methylase (M.SssI) | 200 units 400 units | E2010 E2011 |

